



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115209

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	07/06/2021	14:20
2	M2928-04		1	9	100.00	>212.0	>212.0	07/06/2021	14:45
3	M2928-04DUP		1	9	100.00	>212.0	>212.0	07/06/2021	15:10
4	M2928-07		1	8	100.00	>212.0	>212.0	07/06/2021	15:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB115209

WorkList Name : FP7621 WorkList ID : 150334 Department : Wet-Chemistry Date : 07-06-2021 11:12:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/16/2021	Water	M2928-04	Flash Point	Cool 4 deg C	PSEG03	F51	TOTE-COMP-64-66	07/02/2021	1010A
07/16/2021	Water	M2928-07	Flash Point	Cool 4 deg C	PSEG03	F51	DRUM-1067	07/02/2021	1010A

Date/Time 07/06/21 14:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: MB (1) (2)

Date/Time 07/06/21 15:40
 Raw Sample Received by: MB (D2)
 Raw Sample Relinquished by: KS (WC)