


LB 115214

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5555e-004

C= 1.9806e-002

Rho= 0.9999368

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.030	0.030	38	0.000	38					—
0.2	0.200	0.206	0.006	730	0.0 %	730					3
2.5	2.500	2.453	-0.047	9522	0.0 %	9522					-2
5.0	5.000	4.964	-0.036	19349	0.0 %	19349					-1
7.5	7.500	7.573	0.073	29555	0.0 %	29555					1
10.0	10.000	9.974	-0.026	38951	0.0 %	38951					0

LB115214 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Jul 2021 13:39:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	38	-	0.0000	7470A	PPB	06 Jul 2021 14:02:24	S	Std
0.2 - 1	50.2	730	-	0.2000	7470A	PPB	06 Jul 2021 14:04:41	S	Std
2.5 - 1	52.5	9522	-	2.5000	7470A	PPB	06 Jul 2021 14:06:58	S	Std
5.0 - 1	55	19349	-	5.0000	7470A	PPB	06 Jul 2021 14:09:15	S	Std
7.5 - 1	57.5	29555	-	7.5000	7470A	PPB	06 Jul 2021 14:11:30	S	Std
10.0 - 1	510	38951	-	10.0000	7470A	PPB	06 Jul 2021 14:13:46	S	Std
ICV90 - 1	ICV90	15341	3.9402	-7470A	PPB	06 Jul 2021 14:16:33	U	SMPL	
ICB90 - 1	ICB90	-60	0.0045	-7470A	PPB	06 Jul 2021 14:18:47	U	SMPL	
CCV15 - 1	CCV15	19140	4.9111	-7470A	PPB	06 Jul 2021 14:21:02	U	SMPL	
CCB15 - 1	CCB15	-54	0.0060	-7470A	PPB	06 Jul 2021 14:23:17	U	SMPL	
CRA - 1	CRA	731	0.2066	-7470A	PPB	06 Jul 2021 14:25:32	U	SMPL	
HighStd - 1	HighStd	39080	10.0068	-7470A	PPB	06 Jul 2021 14:27:47	U	SMPL	
ChkStd - 1	ChkStd	27319	7.0012	-7470A	PPB	06 Jul 2021 14:30:02	U	SMPL	
PB137527BL - 1	PBW	-124	-0.0119	-7470A	PPB	06 Jul 2021 14:32:19	U	SMPL	
PB137527BS - 1	LCSW	15477	3.9750	-7470A	PPB	06 Jul 2021 14:34:34	U	SMPL	
M2773-16 - 1	SB-2	1958	0.5202	-7470A	PPB	06 Jul 2021 14:36:51	U	SMPL	
M2773-17 - 1	SB-4	557	0.1621	-7470A	PPB	06 Jul 2021 14:39:07	U	SMPL	
M2773-17DUP - 1	SB-4DUP	550	0.1604	-7470A	PPB	06 Jul 2021 14:41:24	U	SMPL	
M2773-17MS - 1	SB-4MS	16210	4.1623	-7470A	PPB	06 Jul 2021 14:43:41	U	SMPL	
M2773-17MSD - 1	SB-4MSD	15346	3.9415	-7470A	PPB	06 Jul 2021 14:48:35	U	SMPL	
CCV16 - 1	CCV16	18860	4.8395	-7470A	PPB	06 Jul 2021 14:50:50	U	SMPL	
CCB16 - 1	CCB16	-86	-0.0022	-7470A	PPB	06 Jul 2021 14:53:05	U	SMPL	
M2773-18 - 1	SB-5	288	0.0934	-7470A	PPB	06 Jul 2021 14:55:21	U	SMPL	
PB137528BL - 1	PBW	6	0.0213	-7470A	PPB	06 Jul 2021 14:57:36	U	SMPL	
PB137528BS - 1	LCSW	16129	4.1416	-7470A	PPB	06 Jul 2021 14:59:51	U	SMPL	
M2926-03 - 1	TP-1	-69	0.0022	-7470A	PPB	06 Jul 2021 15:02:07	U	SMPL	
M2926-03DUP - 1	TP-1DUP	4	0.0208	-7470A	PPB	06 Jul 2021 15:04:22	U	SMPL	
M2926-03MS - 1	TP-1MS	13915	3.5758	-7470A	PPB	06 Jul 2021 15:06:38	U	SMPL	
M2926-03MSD - 1	TP-1MSD	15409	3.9576	-7470A	PPB	06 Jul 2021 15:08:55	U	SMPL	
M2926-06 - 1	TP-2	-77	0.0001	-7470A	PPB	06 Jul 2021 15:11:11	U	SMPL	
M2926-09 - 1	TP-3	14	0.0234	-7470A	PPB	06 Jul 2021 15:13:28	U	SMPL	
PB137529BL - 1	PBW	6	0.0213	-7470A	PPB	06 Jul 2021 15:15:43	U	SMPL	
CCV17 - 1	CCV17	18819	4.8290	-7470A	PPB	06 Jul 2021 15:17:59	U	SMPL	
CCB17 - 1	CCB17	-75	0.0006	-7470A	PPB	06 Jul 2021 15:20:14	U	SMPL	
PB137529BS - 1	LCSW	14198	3.6481	-7470A	PPB	06 Jul 2021 15:22:30	U	SMPL	
M2928-04 - 1	TOTE-COMP-64-66	-85	-0.0019	-7470A	PPB	06 Jul 2021 15:24:45	U	SMPL	
M2928-04DUP - 1	TOTE-COMP-64-66DUP	20	0.0249	-7470A	PPB	06 Jul 2021 15:27:00	U	SMPL	
M2928-04MS - 1	TOTE-COMP-64-66MS	13635	3.5043	-7470A	PPB	06 Jul 2021 15:29:16	U	SMPL	
M2928-04MSD - 1	TOTE-COMP-64-66MSD	14489	3.7225	-7470A	PPB	06 Jul 2021 15:38:22	U	SMPL	
M2928-07 - 1	DRUM-1067	-77	0.0001	-7470A	PPB	06 Jul 2021 15:40:38	U	SMPL	
CCV18 - 1	CCV18	18591	4.7708	-7470A	PPB	06 Jul 2021 15:42:55	U	SMPL	
CCB18 - 1	CCB18	-83	-0.0014	-7470A	PPB	06 Jul 2021 15:45:11	U	SMPL	
PB137527TB - 1	PB137480TB	121	0.0507	-7470A	PPB	06 Jul 2021 15:47:28	U	SMPL	
PB137528TB - 1	PB137504TB	4	0.0208	-7470A	PPB	06 Jul 2021 15:49:44	U	SMPL	
PB137529TB - 1	PB137506TB	11	0.0226	-7470A	PPB	06 Jul 2021 15:52:00	U	SMPL	
M2773-17LX5 - 1		157	0.0599	-7470A	PPB	06 Jul 2021 15:54:16	U	SMPL	
M2773-17A - 1		14766	3.7933	-7470A	PPB	06 Jul 2021 15:56:31	U	SMPL	
M2926-03LX5 - 1		-78	-0.0001	-7470A	PPB	06 Jul 2021 15:58:46	U	SMPL	
M2926-03A - 1		14730	3.7841	-7470A	PPB	06 Jul 2021 16:01:01	U	SMPL	
M2928-04LX5 - 1		-79	-0.0004	-7470A	PPB	06 Jul 2021 16:03:17	U	SMPL	
M2928-04A - 1		14684	3.7723	-7470A	PPB	06 Jul 2021 16:05:33	U	SMPL	
CCV19 - 1	CCV19	19045	4.8868	-7470A	PPB	06 Jul 2021 16:07:49	U	SMPL	
CCB19 - 1	CCB19	-92	-0.0037	-7470A	PPB	06 Jul 2021 16:10:05	U	SMPL	