



Analytical Summary Report

Analysis Method: SM2120 B

Reviewed By: apatel

Parameter: Color

Supervisor Review By: ketankumar

Run Number: LB115215

pH Meter ID: WC PH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP91853
Calibration std, 5 color unit	WP91854
Calibration std, 10 color unit	WP91855
Calibration std, 15 color unit	WP91856
Calibration std, 20 color unit	WP91857
Calibration std, 25 color unit	WP91858
Calibration std, 30 color unit	WP91859
Calibration std, 40 color unit	WP91860
Calibration std, 50 color unit	WP91861
Calibration std, 60 color unit	WP91862
Calibration std, 70 color unit	WP91863

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB115215BL	LB115215BL	1	5.12	N/A	0.00	07/06/2021	14:25
2	M2945-03	FE-19B-SOLUBLE	10	8.09	N/A	70.00	07/06/2021	14:30
3	M2945-03DUP	FE-19B-SOLUBLEDUP	10	8.09	N/A	70.00	07/06/2021	14:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Color-070621

WorkList ID : 150340

Department : Wet-Chemistry

Date : 07-06-2021 14:14:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/08/2021	Water	M2945-03	☒ Color	Cool 4 deg C	M&MM01	L11	FE-19B-SOLUBLE	07/06/2021	SM2120 B

Date/Time 7/6/21 14:15
 Raw Sample Received by: AB (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 7/6/21 14:50
 Raw Sample Received by: SD (WC)
 Raw Sample Relinquished by: AE (WC)