



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115219

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	07/06/2021	13:25
2	CAL2	1	Water	NA	NA	21.1	7.00	07/06/2021	13:26
3	CAL3	1	Water	NA	NA	21.1	10.02	07/06/2021	13:27
4	ICV	1	Water	NA	NA	21.2	6.99	07/06/2021	13:28
5	CCV1	1	Water	NA	NA	21.1	2.01	07/06/2021	13:30
6	M2948-01	1	Water	NA	NA	24.2	6.81	07/06/2021	13:33
7	M2948-03	1	Water	NA	NA	24.4	6.15	07/06/2021	13:36
8	M2948-03DUP	1	Water	NA	NA	24.5	6.16	07/06/2021	13:37
9	CCV2	1	Water	NA	NA	21.1	12.02	07/06/2021	13:39

20B 115219

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2948 WorkList ID : 150339 Department : Wet-Chemistry Date : 07-06-2021 14:13:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/30/2021	Water	M2948-01	F pH	Cool 4 deg C	LOCK01	L12	MW-3	06/30/2021	SM 4500-H B
06/30/2021	Water	M2948-03	F pH	Cool 4 deg C	LOCK01	L12	MW-1	06/30/2021	SM 4500-H B

Date/Time 07/06/21 13:20

Raw Sample Received by: JP WJC

Raw Sample Relinquished by: JP WJC

Date/Time 07/06/21

Raw Sample Received by: JP WJC

Raw Sample Relinquished by: JP WJC