


LB115220

7471B

INSTRUMENT ID: CV1

 μ Abs.:

54628



Linear

A= 0.0000e+000

B= 1.8335e-004

C= -5.3328e-003

Rho= 0.9999675

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.006	-0.006	-3	0.000	-3					-6
0.2	0.200	0.189	-0.011	1062	0.0 %	1062					0
2.5	2.500	2.506	0.006	13696	0.0 %	13696					0
5.0	5.000	5.050	0.050	27571	0.0 %	27571					-1
7.5	7.500	7.450	-0.050	40660	0.0 %	40660					-1
10.0	10.000	10.011	0.011	54628	0.0 %	54628					0

LB115220 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Jul 2021 16:18:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.0000	7471B	PPB	06 Jul 2021 16:21:58	S	Std
0.2 - 1	50.2	1062	-	0.2000	7471B	PPB	06 Jul 2021 16:24:14	S	Std
2.5 - 1	50.5	13696	-	2.5000	7471B	PPB	06 Jul 2021 16:26:31	S	Std
5.0 - 1	55	27571	-	5.0000	7471B	PPB	06 Jul 2021 16:28:48	S	Std
7.5 - 1	59.5	40660	-	7.5000	7471B	PPB	06 Jul 2021 16:31:04	S	Std
10.0 - 1	510	54628	-	10.0000	7471B	PPB	06 Jul 2021 16:38:25	S	Std
ICV91 - 1	ICV91	21544	3.9448	-7471B	PPB	PPB	06 Jul 2021 16:41:02	U	SMPL
ICB91 - 1	ICB91	-105	-0.0246	-7471B	PPB	PPB	06 Jul 2021 16:43:16	U	SMPL
CCV20 - 1	CCV20	27118	4.9668	-7471B	PPB	PPB	06 Jul 2021 16:45:31	U	SMPL
CCB20 - 1	CCB20	-120	-0.0273	-7471B	PPB	PPB	06 Jul 2021 16:47:46	U	SMPL
CRA - 1	CRA	1143	0.2042	-7471B	PPB	PPB	06 Jul 2021 16:50:02	U	SMPL
HighStd - 1	HighStd	54642	10.0135	-7471B	PPB	PPB	06 Jul 2021 16:52:17	U	SMPL
ChkStd - 1	ChkStd	38045	6.9703	-7471B	PPB	PPB	06 Jul 2021 16:54:32	U	SMPL
PB137531BL - 1	PBS	-188	-0.0398	-7471B	PPB	PPB	06 Jul 2021 16:56:49	U	SMPL
PB137531BS - 1	LCSS	21759	3.9843	-7471B	PPB	PPB	06 Jul 2021 16:59:07	U	SMPL
M2926-01 - 1	TP-1	1195	0.2138	-7471B	PPB	PPB	06 Jul 2021 17:01:23	U	SMPL
M2926-04 - 1	TP-2	941	0.1672	-7471B	PPB	PPB	06 Jul 2021 17:03:40	U	SMPL
M2926-07 - 1	TP-3	729	0.1283	-7471B	PPB	PPB	06 Jul 2021 17:05:56	U	SMPL
M2931-01 - 1	HD-01-070221	1386	0.2488	-7471B	PPB	PPB	06 Jul 2021 17:08:13	U	SMPL
M2932-01 - 1	DRUM-1	1095	0.1954	-7471B	PPB	PPB	06 Jul 2021 17:10:28	U	SMPL
CCV21 - 1	CCV21	27116	4.9665	-7471B	PPB	PPB	06 Jul 2021 17:12:43	U	SMPL
CCB21 - 1	CCB21	-108	-0.0251	-7471B	PPB	PPB	06 Jul 2021 17:14:58	U	SMPL
M2933-01 - 1	1	2433	0.4408	-7471B	PPB	PPB	06 Jul 2021 17:17:13	U	SMPL
M2933-01DUP - 1	1DUP	2354	0.4263	-7471B	PPB	PPB	06 Jul 2021 17:19:28	U	SMPL
M2933-01MS - 1	1MS	24905	4.5611	-7471B	PPB	PPB	06 Jul 2021 17:21:44	U	SMPL
M2933-01MSD - 1	1MSD	24901	4.5604	-7471B	PPB	PPB	06 Jul 2021 17:23:59	U	SMPL
M2933-02 - 1	2	4579	0.8342	-7471B	PPB	PPB	06 Jul 2021 17:26:15	U	SMPL
M2933-03 - 1	3	1867	0.3370	-7471B	PPB	PPB	06 Jul 2021 17:28:31	U	SMPL
M2933-04 - 1	4	3565	0.6483	-7471B	PPB	PPB	06 Jul 2021 17:30:47	U	SMPL
M2933-05 - 1	5	3049	0.5537	-7471B	PPB	PPB	06 Jul 2021 17:33:04	U	SMPL
M2936-01 - 1	TR-05-070221	1708	0.3078	-7471B	PPB	PPB	06 Jul 2021 17:35:21	U	SMPL
M2938-01 - 1	OK-01-070221	1850	0.3339	-7471B	PPB	PPB	06 Jul 2021 17:37:36	U	SMPL
CCV22 - 1	CCV22	27136	4.9701	-7471B	PPB	PPB	06 Jul 2021 17:39:51	U	SMPL
CCB22 - 1	CCB22	-93	-0.0224	-7471B	PPB	PPB	06 Jul 2021 17:42:07	U	SMPL
M2933-01LX5 - 1		453	0.0777	-7471B	PPB	PPB	06 Jul 2021 17:44:22	U	SMPL
M2933-01A - 1		24080	4.4098	-7471B	PPB	PPB	06 Jul 2021 17:46:37	U	SMPL
CCV23 - 1	CCV23	26684	4.8873	-7471B	PPB	PPB	06 Jul 2021 17:48:53	U	SMPL
CCB23 - 1	CCB23	-116	-0.0266	-7471B	PPB	PPB	06 Jul 2021 17:51:09	U	SMPL