



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115222

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/06/2021	16:25
2	CAL2	1	Water	NA	NA	21.2	6.99	07/06/2021	16:26
3	CAL3	1	Water	NA	NA	21.2	10.02	07/06/2021	16:27
4	ICV	1	Water	NA	NA	21.3	7.01	07/06/2021	16:29
5	CCV1	1	Water	NA	NA	21.3	2.01	07/06/2021	16:30
6	M2944-01	1	Solid	20.02	20	24.4	6.64	07/06/2021	16:32
7	M2944-01DUP	1	Solid	20.03	20	24.5	6.65	07/06/2021	16:33
8	CCV2	1	Water	NA	NA	21.4	12.02	07/06/2021	16:35

2B115222

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2944 WorkList ID : 150345 Department : Wet-Chemistry Date : 07-06-2021 16:19:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/06/2021	Solid	M2944-01	pH	Cool 4 deg C	PSEG05	L12	BU-03-070621	07/06/2021	9045D

Date/Time 07/06/21 16:20
Raw Sample Received by: JP CWC
Raw Sample Relinquished by: JP CWC

Date/Time 07/06/21 17:30
Raw Sample Received by: JP CWC
Raw Sample Relinquished by: JP CWC