


LB115230

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4827e-004

C= 2.1724e-002

Rho= 0.9997924

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.000	-0.000	-88	0.000	-88					-9
0.2	0.200	0.183	-0.017	651	0.0 %	651					-1
2.5	2.500	2.482	-0.018	9908	0.0 %	9908					0
5.0	5.000	4.993	-0.007	20025	0.0 %	20025					2
7.5	7.500	7.648	0.148	30719	0.0 %	30719					-1
10.0	10.000	9.894	-0.106	39764	0.0 %	39764					

LB115230 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Jul 2021 14:01:52					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-88	-	0.0000	7470A	PPB	07 Jul 2021 14:11:32	S	Std
0.2 - 1	50.2	651	-	0.2000	7470A	PPB	07 Jul 2021 14:13:49	S	Std
2.5 - 1	52.5	9908	-	2.5000	7470A	PPB	07 Jul 2021 14:16:06	S	Std
5.0 - 1	55	20025	-	5.0000	7470A	PPB	07 Jul 2021 14:18:23	S	Std
7.5 - 1	57.5	30719	-	7.5000	7470A	PPB	07 Jul 2021 14:20:39	S	Std
10.0 - 1	510	39764	-	10.0000	7470A	PPB	07 Jul 2021 14:22:55	S	Std
ICV92 - 1	ICV92	15984	3.9900	-	7470A	PPB	07 Jul 2021 14:25:51	U	SMPL
ICB92 - 1	ICB92	-50	0.0093	-	7470A	PPB	07 Jul 2021 14:28:05	U	SMPL
CCV24 - 1	CCV24	19628	4.8947	-	7470A	PPB	07 Jul 2021 14:32:53	U	SMPL
CCB24 - 1	CCB24	-68	0.0048	-	7470A	PPB	07 Jul 2021 14:35:08	U	SMPL
CRA - 1	CRA	701	0.1958	-	7470A	PPB	07 Jul 2021 14:37:23	U	SMPL
HighStd - 1	HighStd	38617	9.6090	-	7470A	PPB	07 Jul 2021 14:39:37	U	SMPL
ChkStd - 1	ChkStd	25963	6.4675	-	7470A	PPB	07 Jul 2021 14:41:52	U	SMPL
PB137556BL - 1	PBW	-16	0.0178	-	7470A	PPB	07 Jul 2021 14:46:46	U	SMPL
PB137556BS - 1	LCSW	15812	3.9473	-	7470A	PPB	07 Jul 2021 14:52:15	U	SMPL
M2949-01 - 1	POLY-149TH-ST-STATION	-35	0.0130	-	7470A	PPB	07 Jul 2021 14:54:31	U	SMPL
M2949-01DUP - 1	POLY-149TH-ST-STATIONDUP	16	0.0257	-	7470A	PPB	07 Jul 2021 14:56:47	U	SMPL
M2949-01MS - 1	POLY-149TH-ST-STATIONMS	14294	3.5704	-	7470A	PPB	07 Jul 2021 15:01:35	U	SMPL
M2949-01MSD - 1	POLY-149TH-ST-STATIONMSD	14543	3.6323	-	7470A	PPB	07 Jul 2021 15:03:52	U	SMPL
M2949-02 - 1	PPE-149TH-ST-STATION	-45	0.0106	-	7470A	PPB	07 Jul 2021 15:06:08	U	SMPL
CCV25 - 1	CCV25	19074	4.7572	-	7470A	PPB	07 Jul 2021 15:08:22	U	SMPL
CCB25 - 1	CCB25	-79	0.0021	-	7470A	PPB	07 Jul 2021 15:10:38	U	SMPL
PB137556TB - 1	PB137543TB	-24	0.0158	-	7470A	PPB	07 Jul 2021 15:12:56	U	SMPL
M2949-01LX5 - 1		-7	0.0200	-	7470A	PPB	07 Jul 2021 15:15:11	U	SMPL
M2949-01A - 1		15235	3.8041	-	7470A	PPB	07 Jul 2021 15:17:26	U	SMPL
CCV26 - 1	CCV26	18871	4.7068	-	7470A	PPB	07 Jul 2021 15:19:42	U	SMPL
CCB26 - 1	CCB26	-77	0.0026	-	7470A	PPB	07 Jul 2021 15:21:57	U	SMPL