



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB115237

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP91895
Turbidity Calibration std, 0NTU	WP91894
Turbidity Calibration std, 20NTU	WP91890
Turbidity Calibration std, 40NTU	WP91889
Turbidity Calibration std, 1NTU	WP91893
Turbidity Calibration std, 5NTU	WP91892
Turbidity LOQ std, 1.0NTU	WP91896
Turbidity Calibration - CCV std, 10 NTU	WP91891
10 NTU Standard 500 ml	W2752

Intercept: 0.1874

Slope: 0.9907

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.181	-0.01		07/07/2021	12:31
2	CAL2	1	1	1.170	0.99	0.8	07/07/2021	12:33
3	CAL3	5	1	5.260	5.12	-2.4	07/07/2021	12:35
4	CAL4	10	1	9.910	9.81	1.9	07/07/2021	12:37
5	CAL5	20	1	20.100	20.10	-0.5	07/07/2021	12:39
6	CAL6	40	1	39.800	39.98	0	07/07/2021	12:42



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB115237

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.200	10.107	07/07/2021	12:49
2	ICB		1	0.185	-0.002	07/07/2021	12:52
3	CCV1	10	1	10.100	10.006	07/07/2021	12:54
4	CCB1		1	0.188	0.001	07/07/2021	12:59
5	LB115237BL		1	0.189	0.002	07/07/2021	13:02
6	M2940-10		1	0.535	0.351	07/07/2021	13:04
7	M2940-11		1	1.150	0.972	07/07/2021	13:08
8	M2948-01		1	0.433	0.248	07/07/2021	13:11
9	M2948-03		1	0.309	0.123	07/07/2021	13:13
10	M2948-03DUP		1	0.312	0.126	07/07/2021	13:16
11	CCV2	10	1	10.100	10.006	07/07/2021	13:19
12	CCB2		1	0.191	0.004	07/07/2021	13:21

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-070721 WorkList ID : 150370 Department : Wet-Chemistry Date : 07-07-2021 12:06:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/08/2021	Water	M2940-07	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOD-MDL-WATER-01-QT	07/06/2021	180.1
07/08/2021	Water	M2940-08	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOQ-WATER-02-QT3-20	07/06/2021	180.1
07/08/2021	Water	M2940-10	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOD-MDL-WATER-04-QT	07/06/2021	SM2130 B
07/08/2021	Water	M2940-11	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOQ-WATER-05-QT3-20	07/06/2021	SM2130 B
07/02/2021	Water	M2948-01	Turbidity	Cool 4 deg C	LOCK01	L12	MW-3	06/30/2021	SM2130 B
07/02/2021	Water	M2948-03	Turbidity	Cool 4 deg C	LOCK01	L12	MW-1	06/30/2021	SM2130 B

Date/Time 7/7/21 12:10
Raw Sample Received by: AO (WCC)
Raw Sample Relinquished by: SD (WCC)

Date/Time 7/7/21 13:20
Raw Sample Received by: SP (WCC)
Raw Sample Relinquished by: AO (WCC)