

Analytical Summary Report

Analysis Method: 180.1

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB115239

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP91895
Turbidity Calibration std, 0NTU	WP91894
Turbidity Calibration std, 20NTU	WP91890
Turbidity Calibration std, 40NTU	WP91889
Turbidity Calibration std, 1NTU	WP91893
Turbidity Calibration std, 5NTU	WP91892
Turbidity LOQ std, 1.0NTU	WP91896
Turbidity Calibration - CCV std, 10 NTU	WP91891
10 NTU Standard 500 ml	W2752

Intercept: 0.1874

Slope: 0.9907

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.181	-0.01		07/07/2021	12:31
2	CAL2	1	1	1.170	0.99	0.8	07/07/2021	12:33
3	CAL3	5	1	5.260	5.12	-2.4	07/07/2021	12:35
4	CAL4	10	1	9.910	9.81	1.9	07/07/2021	12:37
5	CAL5	20	1	20.100	20.10	-0.5	07/07/2021	12:39
6	CAL6	40	1	39.800	39.98	0	07/07/2021	12:42



Analytical Summary Report

Analysis Method: 180.1

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB115239

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.200	10.107	07/07/2021	12:49
2	ICB		1	0.185	-0.002	07/07/2021	12:52
3	CCV1	10	1	10.100	10.006	07/07/2021	12:54
4	CCB1		1	0.188	0.001	07/07/2021	12:59
5	LB115239BL		1	0.189	0.002	07/07/2021	13:02
6	M2940-07		1	0.535	0.351	07/07/2021	13:04
7	M2940-08		1	1.150	0.972	07/07/2021	13:08
8	CCV2	10	1	10.100	10.006	07/07/2021	13:19
9	CCB2		1	0.191	0.004	07/07/2021	13:21