



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115248

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.5	4.01	07/08/2021	08:15
2	CAL2	1	Water	NA	NA	21.5	7.01	07/08/2021	08:16
3	CAL3	1	Water	NA	NA	21.5	10.03	07/08/2021	08:17
4	ICV	1	Water	NA	NA	21.6	6.99	07/08/2021	08:19
5	CCV1	1	Water	NA	NA	21.5	2.01	07/08/2021	08:20
6	M2967-01	1	Solid	20.02	20	24.9	7.31	07/08/2021	08:22
7	M2967-01DUP	1	Solid	20.03	20	25.0	7.32	07/08/2021	08:23
8	M2967-05	1	Solid	20.02	20	25.0	7.53	07/08/2021	08:25
9	M2967-09	1	Solid	20.03	20	24.9	8.58	07/08/2021	08:26
10	CCV2	1	Water	NA	NA	21.5	12.02	07/08/2021	08:29

VB115248

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2967 WorkList ID : 150384 Department : Wet-Chemistry Date : 07-08-2021 08:09:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/07/2021	Solid	M2967-01	Corrosivity	Cool 4 deg C	PSEG03	L12	1-5ft-GRACE-AVE-TP-1	07/07/2021	9045D
07/07/2021	Solid	M2967-05	Corrosivity	Cool 4 deg C	PSEG03	L12	1-8FT-STEFANIC-AVE-TF	07/07/2021	9045D
07/07/2021	Solid	M2967-09	Corrosivity	Cool 4 deg C	PSEG03	L12	1-5FT-CADMUS-AVE-TP-	07/07/2021	9045D

Date/Time 07/08/21 08:10
 Raw Sample Received by: CP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 07/08/21 11:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: CP (WC)