



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115249

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.5	4.01	07/08/2021	08:15
2	CAL2	1	Water	NA	NA	21.5	7.01	07/08/2021	08:16
3	CAL3	1	Water	NA	NA	21.5	10.03	07/08/2021	08:17
4	ICV	1	Water	NA	NA	21.6	6.99	07/08/2021	08:19
5	CCV1	1	Water	NA	NA	21.5	2.01	07/08/2021	08:20
6	M2966-01	1	Solid	20.04	20	24.4	9.06	07/08/2021	08:32
7	M2966-01DUP	1	Solid	20.03	20	24.5	9.07	07/08/2021	08:33
8	M2966-02	1	Solid	20.02	20	24.3	8.97	07/08/2021	08:34
9	M2966-03	1	Solid	20.01	20	24.0	8.83	07/08/2021	08:35
10	M2966-04	1	Solid	20.03	20	23.7	9.45	07/08/2021	08:36
11	CCV2	1	Water	NA	NA	21.5	12.02	07/08/2021	08:39

115249

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2979 WorkList ID : 150383 Department : Wet-Chemistry Date : 07-08-2021 08:08:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/07/2021	Solid	M2966-01	pH	Cool 4 deg C	GEIC06	L12	(WINDROW-LEFT-OF-EN	07/07/2021	9045D
07/07/2021	Solid	M2966-02	pH	Cool 4 deg C	GEIC06	L12	(MIDDLE-WINDROW)	07/07/2021	9045D
07/07/2021	Solid	M2966-03	pH	Cool 4 deg C	GEIC06	L12	(RIGHT-WINDROW)	07/07/2021	9045D
07/07/2021	Solid	M2966-04	pH	Cool 4 deg C	GEIC06	L12	TOP-SOIL-PILE	07/07/2021	9045D

Date/Time 07/08/21 08:11
Raw Sample Received by: JB (WC)
Raw Sample Relinquished by: CP SM

Date/Time 07/08/21 11:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JB (WC)