



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115253

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/08/2021	12:33
2	CAL2	1	Water	NA	NA	21.3	7.00	07/08/2021	12:34
3	CAL3	1	Water	NA	NA	21.3	10.02	07/08/2021	12:35
4	ICV	1	Water	NA	NA	21.4	7.00	07/08/2021	12:36
5	CCV1	1	Water	NA	NA	21.4	2.01	07/08/2021	12:38
6	M2972-01	1	Water	NA	NA	24.4	6.63	07/08/2021	12:40
7	M2974-01	1	Water	NA	NA	24.4	5.51	07/08/2021	12:41
8	M2974-02	1	Water	NA	NA	24.5	5.24	07/08/2021	12:43
9	M2975-01	1	Water	NA	NA	21.1	6.76	07/08/2021	12:45
10	M2975-03	1	Water	NA	NA	20.9	6.87	07/08/2021	12:49
11	M2975-03DUP	1	Water	NA	NA	21.1	6.88	07/08/2021	12:50
12	CCV2	1	Water	NA	NA	21.4	12.02	07/08/2021	13:00

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2975 WorkList ID : 150403 Department : Wet-Chemistry Date : 07-08-2021 12:27:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/06/2021	Water	M2972-01	pH	Cool 4 deg C	STER08	L11	POOL-WATER	07/06/2021	9040C
07/07/2021	Water	M2974-01	pH	Cool 4 deg C	STER16	L11	MAIN-CELL	07/07/2021	9040C
07/07/2021	Water	M2974-02	pH	Cool 4 deg C	STER16	L11	EXCELL	07/07/2021	9040C
07/07/2021	Water	M2975-01	pH	Cool 4 deg C	APTI01	L11	WGGM-21-WW-TSF-028	07/07/2021	9040C
07/07/2021	Water	M2975-03	pH	Cool 4 deg C	APTI01	L11	WGGM-21-WW-TSF-029	07/07/2021	9040C

Date/Time 07/08/21 12:30
 Raw Sample Received by: JP (wc)
 Raw Sample Relinquished by: JP (wc)

Date/Time 07/08/21 14:00
 Raw Sample Received by: JP (wc)
 Raw Sample Relinquished by: JP (wc)