


LB115261

7470A INSTRUMENT ID: CV 1

 μ Abs.:

37290



Linear

A= 0.0000e+000

B= 2.6848e-004

C= -2.5979e-002

Rho= 0.9999954

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.010	-0.010	58	0.000	58					-
0.2	0.200	0.203	0.003	854	0.0 %	854					2
2.5	2.500	2.495	-0.005	9389	0.0 %	9389					0
5.0	5.000	5.018	0.018	18787	0.0 %	18787					0
7.5	7.500	7.509	0.009	28064	0.0 %	28064					0
10.0	10.000	9.986	-0.014	37290	0.0 %	37290					0

LB115261 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Jul 2021 15:37:02

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	58	-	0.00007470A	PPB	08 Jul 2021 15:37:45	S	Std
0.2 - 1	50.2	854	-	0.20007470A	PPB	08 Jul 2021 15:40:01	S	Std
2.5 - 1	52.5	9389	-	2.50007470A	PPB	08 Jul 2021 15:42:18	S	Std
5.0 - 1	15	18787	-	5.00007470A	PPB	08 Jul 2021 15:44:35	S	Std
7.5 - 1	57.5	28064	-	7.50007470A	PPB	08 Jul 2021 15:46:51	S	Std
10.0 - 1	51V	37290	-	10.00007470A	PPB	08 Jul 2021 15:54:55	S	Std
ICV95 - 1	ICV95	15171	4.0471	-7470A	PPB	08 Jul 2021 16:03:50	U	SMPL
ICB95 - 1	ICB95	-66	-0.0437	-7470A	PPB	08 Jul 2021 16:06:05	U	SMPL
CCV34 - 1	CCV34	17667	4.7173	-7470A	PPB	08 Jul 2021 16:08:20	U	SMPL
CCB34 - 1	CCB34	-66	-0.0437	-7470A	PPB	08 Jul 2021 16:10:35	U	SMPL
CRA - 1	CRA	775	0.1821	-7470A	PPB	08 Jul 2021 16:12:50	U	SMPL
HighStd - 1	HighStd	35307	9.4533	-7470A	PPB	08 Jul 2021 16:15:05	U	SMPL
ChkStd - 1	ChkStd	24600	6.5787	-7470A	PPB	08 Jul 2021 16:17:20	U	SMPL
PB137579BL - 1	PBW	-30	-0.0340	-7470A	PPB	08 Jul 2021 16:19:37	U	SMPL
PB137579BS - 1	LCSW	13778	3.6732	-7470A	PPB	08 Jul 2021 16:24:25	U	SMPL
M2965-03 - 1	RO-ETGI-345	-6	-0.0276	-7470A	PPB	08 Jul 2021 16:26:40	U	SMPL
M2967-02 - 1	1-5ft-GRACE-AVE-TP-1	60	-0.0099	-7470A	PPB	08 Jul 2021 16:28:56	U	SMPL
M2967-06 - 1	1-8FT-STEFANIC-AVE-TP-2	62	-0.0093	-7470A	PPB	08 Jul 2021 16:31:14	U	SMPL
M2967-10 - 1	1-5FT-CADMUS-AVE-TP-3	26	-0.0190	-7470A	PPB	08 Jul 2021 16:33:31	U	SMPL
M2967-10DUP - 1	1-5FT-CADMUS-AVE-TP-3DUP	21	-0.0203	-7470A	PPB	08 Jul 2021 16:35:47	U	SMPL
CCV35 - 1	CCV35	18572	4.9602	-7470A	PPB	08 Jul 2021 16:40:32	U	SMPL
CCB35 - 1	CCB35	-163	-0.0697	-7470A	PPB	08 Jul 2021 16:42:48	U	SMPL
M2967-10MS - 1	1-5FT-CADMUS-AVE-TP-3MS	15003	4.0020	-7470A	PPB	08 Jul 2021 16:45:02	U	SMPL
M2967-10MSD - 1	1-5FT-CADMUS-AVE-TP-3MSD	14753	3.9349	-7470A	PPB	08 Jul 2021 16:47:17	U	SMPL
PB137579TB - 1	PB137560TB	27	-0.0187	-7470A	PPB	08 Jul 2021 16:49:32	U	SMPL
M2967-10LX5 - 1		-4	-0.0271	-7470A	PPB	08 Jul 2021 16:51:48	U	SMPL
M2967-10A - 1		15239	4.0654	-7470A	PPB	08 Jul 2021 16:54:04	U	SMPL
CCV36 - 1	CCV36	19118	5.1068	-7470A	PPB	08 Jul 2021 16:56:20	U	SMPL
CCB36 - 1	CCB36	-138	-0.0630	-7470A	PPB	08 Jul 2021 16:58:36	U	SMPL