



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115264

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/08/2021	16:25
2	CAL2	1	Water	NA	NA	21.3	7.01	07/08/2021	16:26
3	CAL3	1	Water	NA	NA	21.3	10.03	07/08/2021	16:27
4	ICV	1	Water	NA	NA	21.4	6.99	07/08/2021	16:30
5	CCV1	1	Water	NA	NA	21.4	2.01	07/08/2021	16:32
6	M2969-02	1	Solid	20.02	20	24.2	6.56	07/08/2021	16:35
7	M2969-04	1	Solid	20.03	20	24.3	7.24	07/08/2021	16:36
8	M2969-06	1	Solid	20.05	20	24.2	7.52	07/08/2021	16:39
9	M2969-08	1	Solid	20.02	20	24.3	8.00	07/08/2021	16:40
10	M2969-10	1	Solid	20.03	20	24.0	7.72	07/08/2021	16:42
11	M2969-12	1	Solid	20.04	20	24.2	7.62	07/08/2021	16:44
12	M2969-14	1	Solid	20.01	20	23.8	11.14	07/08/2021	16:45
13	M2969-16	1	Solid	20.03	20	24.2	8.85	07/08/2021	16:47
14	M2984-01	1	Solid	20.04	20	24.7	6.55	07/08/2021	16:50
15	M2984-01DUP	1	Solid	20.03	20	24.8	6.56	07/08/2021	16:51
16	CCV2	1	Water	NA	NA	21.4	12.02	07/08/2021	16:52
17	M2984-03	1	Solid	20.04	20	24.9	9.14	07/08/2021	17:00
18	M2990-01	1	Solid	20.02	20	25.1	6.96	07/08/2021	17:02
19	M2990-04	1	Solid	20.03	20	25.1	7.19	07/08/2021	17:05
20	M2990-07	1	Solid	20.04	20	24.9	7.39	07/08/2021	17:10

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M2990-10	1	Solid	20.02	20	24.9	7.39	07/08/2021	17:11
22	M2990-13	1	Solid	20.03	20	24.7	6.61	07/08/2021	17:13
23	M2990-16	1	Solid	20.04	20	25.1	7.68	07/08/2021	17:15
24	CCV3	1	Water	NA	NA	21.4	2.01	07/08/2021	17:16

VB 115264

WORKLIST(Hardcopy Internal Chain)

Date : 07-08-2021 13:18:25

Department : Wet-Chemistry

WorkList ID : 150408

WorkList Name : CORROSIONITY M2969

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/06/2021	Solid	M2969-02	Corrosivity	Cool 4 deg C	loui01	L11	SB01	07/06/2021	9045D
07/06/2021	Solid	M2969-04	Corrosivity	Cool 4 deg C	loui01	L11	SB02	07/06/2021	9045D
07/06/2021	Solid	M2969-06	Corrosivity	Cool 4 deg C	loui01	L11	SB03	07/06/2021	9045D
07/07/2021	Solid	M2969-08	Corrosivity	Cool 4 deg C	loui01	L11	SB04	07/07/2021	9045D
07/07/2021	Solid	M2969-10	Corrosivity	Cool 4 deg C	loui01	L11	SB05	07/07/2021	9045D
07/07/2021	Solid	M2969-12	Corrosivity	Cool 4 deg C	loui01	L11	SB06	07/07/2021	9045D
07/06/2021	Solid	M2969-14	Corrosivity	Cool 4 deg C	loui01	L11	SB07	07/06/2021	9045D
07/06/2021	Solid	M2969-16	Corrosivity	Cool 4 deg C	loui01	L11	SB08	07/06/2021	9045D
07/08/2021	Solid	M2984-01	Corrosivity	Cool 4 deg C	PSEG03	L13	72-11940	07/08/2021	9045D
07/08/2021	Solid	M2984-03	Corrosivity	Cool 4 deg C	PSEG03	L13	VNJ-228	07/08/2021	9045D
07/08/2021	Solid	M2990-01	Corrosivity	Cool 4 deg C	PSEG03	L13	TP-D	07/08/2021	9045D
07/08/2021	Solid	M2990-04	Corrosivity	Cool 4 deg C	PSEG03	L13	TP-G	07/08/2021	9045D
07/08/2021	Solid	M2990-07	Corrosivity	Cool 4 deg C	PSEG03	L13	TP-K	07/08/2021	9045D
07/08/2021	Solid	M2990-10	Corrosivity	Cool 4 deg C	PSEG03	L13	TP-I	07/08/2021	9045D
07/08/2021	Solid	M2990-13	Corrosivity	Cool 4 deg C	PSEG03	L13	TP-J	07/08/2021	9045D
07/08/2021	Solid	M2990-16	Corrosivity	Cool 4 deg C	PSEG03	L13	TP-E	07/08/2021	9045D

Date/Time 07/08/21 15:25
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 07/08/21 18:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: SP (WC)