



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115265

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/08/2021	16:25
2	CAL2	1	Water	NA	NA	21.3	7.01	07/08/2021	16:26
3	CAL3	1	Water	NA	NA	21.3	10.03	07/08/2021	16:27
4	ICV	1	Water	NA	NA	21.4	6.99	07/08/2021	16:30
5	CCV1	1	Water	NA	NA	21.4	2.01	07/08/2021	16:32
6	M2984-01	1	Solid	20.04	20	24.7	6.55	07/08/2021	16:50
7	M2984-01DUP	1	Solid	20.03	20	24.8	6.56	07/08/2021	16:51
8	M2984-03	1	Solid	20.04	20	24.9	9.14	07/08/2021	17:00
9	M2987-01	1	Solid	20.02	20	24.9	8.28	07/08/2021	17:04
10	CCV2	1	Water	NA	NA	21.4	12.03	07/08/2021	17:17

VB 115263

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2969 WorkList ID : 150409 Department : Wet-Chemistry Date : 07-08-2021 13:19:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/08/2021	Solid	M2984-01	pH	Cool 4 deg C	PSEG03	L13	72-11940	07/08/2021	9045D
07/08/2021	Solid	M2984-03	pH	Cool 4 deg C	PSEG03	L13	VNJ-228	07/08/2021	9045D
07/08/2021	Solid	M2987-01	pH	Cool 4 deg C	PSEG03	L13	CHRT-24368	07/08/2021	9045D

Date/Time 07/08/21 15:25
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JS (WC)

Date/Time 07/08/21 18:00
Raw Sample Received by: JS (WC)
Raw Sample Relinquished by: JP (WC)