



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115266

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/08/2021	16:25
2	CAL2	1	Water	NA	NA	21.3	7.01	07/08/2021	16:26
3	CAL3	1	Water	NA	NA	21.3	10.03	07/08/2021	16:27
4	ICV	1	Water	NA	NA	21.4	6.99	07/08/2021	16:30
5	CCV1	1	Water	NA	NA	21.4	2.01	07/08/2021	16:32
6	M2992-01	1	Water	NA	NA	23.5	6.78	07/08/2021	17:25
7	M2992-01DUP	1	Water	NA	NA	23.6	6.79	07/08/2021	17:26
8	M2992-03	1	Water	NA	NA	22.0	4.60	07/08/2021	17:28
9	M2992-05	1	Water	NA	NA	22.5	5.83	07/08/2021	17:29
10	CCV2	1	Water	NA	NA	21.5	12.03	07/08/2021	17:30

VB 115266

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2992 WorkList ID : 150417 Department : Wet-Chemistry Date : 07-08-2021 16:30:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/08/2021	Water	M2992-01	pH	Cool 4 deg C	PSEG03	L11	33287	07/08/2021	9040C
07/08/2021	Water	M2992-03	pH	Cool 4 deg C	PSEG03	L11	33262	07/08/2021	9040C
07/08/2021	Water	M2992-05	pH	Cool 4 deg C	PSEG03	L11	33263	07/08/2021	9040C

Date/Time 07/08/21 16:35
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 07/08/21 18:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JP (WC)