

Analysis Method: 9034

Parameter: Sulfide

Run Number: LB115270

ANALYST: ketankumar

SUPERVISOR REVIEW BY: apatel

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP90938
IODINE SOLUTION .025N 1L	W2763
Starch Solution, 4L	W2703

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB137588BL		1	50	50	5.00	0.00	4.90	4.90	0.10	0.00	0.00	07/08/2021	15:35
2	PB137588BS	25	1	50	50	5.00	0.00	2.22	2.22	2.78	2.68	21.44	07/08/2021	15:38
3	M2940-07		1	50	50	5.00	0.00	4.84	4.84	0.16	0.06	0.48	07/08/2021	15:40
4	M2940-08		1	50	50	5.00	0.00	4.80	4.80	0.20	0.10	0.80	07/08/2021	15:42

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume