

LB115272


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

31005



A= 0.0000e+000

B= 3.2555e-004

C= -5.0335e-002

Rho= 0.9997259

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Q/D
0.0	0.000	-0.015	-0.015	110	0.000	110					-
0.05	0.050	0.041	-0.009	280	0.0 %	280	0				-18
0.2	0.200	0.192	-0.008	744	0.0 %	744					-4
2.5	2.500	2.504	0.004	7846	0.0 %	7846					0
5.0	5.000	5.151	0.151	15978	0.0 %	15978					3
7.5	7.500	7.333	-0.167	22681	0.0 %	22681					-2
10.0	10.000	10.043	0.043	31005	0.0 %	31005					0

LB115272 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 09 Jul 2021 11:56:02				
Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	110	-	0.0000	245.1	PPB	09 Jul 2021 12:00:45	S Std
0.05 - 1	50.05	280	-	0.0500	245.1	PPB	09 Jul 2021 12:03:01	S Std
0.2 - 1	50.2	744	-	0.2000	245.1	PPB	09 Jul 2021 12:05:18	S Std
2.5 - 1	52.5	7846	-	2.5000	245.1	PPB	09 Jul 2021 12:07:34	S Std
5.0 - 1	55	15978	-	5.0000	245.1	PPB	09 Jul 2021 12:09:50	S Std
7.5 - 1	57.5	22681	-	7.5000	245.1	PPB	09 Jul 2021 12:12:06	S Std
10.0 - 1	59.0	31005	-	10.0000	245.1	PPB	09 Jul 2021 12:14:22	S Std
ICV96 - 1	ICV96	12272	3.9448	-245.1		PPB	09 Jul 2021 12:17:16	U SMPL
ICB96 - 1	ICB96	97	-0.0188	-245.1		PPB	09 Jul 2021 12:19:30	U SMPL
CCV37 - 1	CCV37	16048	5.1740	-245.1		PPB	09 Jul 2021 12:21:45	U SMPL
CCB37 - 1	CCB37	96	-0.0191	-245.1		PPB	09 Jul 2021 12:24:00	U SMPL
CRA - 1	CRA	318	0.0532	-245.1		PPB	09 Jul 2021 12:37:24	U SMPL
HighStd - 1	HighStd	29423	9.5282	-245.1		PPB	09 Jul 2021 12:44:45	U SMPL
ChkStd - 1	ChkStd	22699	7.3393	-245.1		PPB	09 Jul 2021 12:47:00	U SMPL
PB137604BS - 1	LCSW	12017	3.8618	-245.1		PPB	09 Jul 2021 13:00:29	U SMPL
M2940-10 - 1	LOD-MDL-WATER-04-QT3-2021	343	0.0613	-245.1		PPB	09 Jul 2021 13:08:53	U SMPL
M2940-11 - 1	LOQ-WATER-05-QT3-2021	654	0.1626	-245.1		PPB	09 Jul 2021 13:11:10	U SMPL
M2955-02 - 1	COMP	693	0.1753	-245.1		PPB	09 Jul 2021 13:13:26	U SMPL
M2955-02DUP - 1	COMPDUP	630	0.1548	-245.1		PPB	09 Jul 2021 13:15:43	U SMPL
M2955-02MS - 1	COMPMS	10971	3.5212	-245.1		PPB	09 Jul 2021 13:17:58	U SMPL
CCV38 - 1	CCV38	16888	5.4475	-245.1		PPB	09 Jul 2021 13:22:49	U SMPL
CCB38 - 1	CCB38	148	-0.0022	-245.1		PPB	09 Jul 2021 13:25:03	U SMPL
M2955-02MSD - 1	COMPMSD	11392	3.6583	-245.1		PPB	09 Jul 2021 13:32:24	U SMPL
PB137604BL - 1	PBW	12	-0.0464	-245.1		PPB	09 Jul 2021 13:34:38	U SMPL
M2955-02LX5 - 1		319	0.0535	-245.1		PPB	09 Jul 2021 13:36:55	U SMPL
M2955-02A - 1		14071	4.5304	-245.1		PPB	09 Jul 2021 13:39:11	U SMPL
CCV39 - 1	CCV39	15295	4.9289	-245.1		PPB	09 Jul 2021 13:41:26	U SMPL
CCB39 - 1	CCB39	136	-0.0061	-245.1		PPB	09 Jul 2021 13:51:13	U SMPL