



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115275

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Falshpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.4	07/14/2021	10:30
2	M2992-01		1	8	100.00	>212.0	>212.0	07/14/2021	10:46
3	M2992-01DUP		1	9	100.00	>212.0	>212.0	07/14/2021	10:59
4	M2992-03		1	8	100.00	>212.0	>212.0	07/14/2021	11:12
5	M2992-05		1	7	100.00	>212.0	>212.0	07/14/2021	11:30
6	M2999-01		1	8	100.00	>212.0	>212.0	07/14/2021	11:49
7	M3027-03		1	7	100.00	>212.0	>212.0	07/14/2021	12:10
8	M3027-05		1	9	100.00	>212.0	>212.0	07/14/2021	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06