



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115277

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/09/2021	13:20
2	CAL2	1	Water	NA	NA	21.2	7.00	07/09/2021	13:21
3	CAL3	1	Water	NA	NA	21.2	10.02	07/09/2021	13:22
4	ICV	1	Water	NA	NA	21.3	6.99	07/09/2021	13:25
5	CCV1	1	Water	NA	NA	21.2	2.01	07/09/2021	13:26
6	M2999-01	1	Water	NA	NA	24.0	8.60	07/09/2021	13:29
7	M2999-01DUP	1	Water	NA	NA	24.1	8.61	07/09/2021	13:30
8	CCV2	1	Water	NA	NA	21.3	12.02	07/09/2021	13:31

187 115277

WORKLIST(Hardcopy Internal Chain)

Date : 07-09-2021 13:11:09

Department : Wet-Chemistry

WorkList ID : 150452

WorkList Name : PH M2999

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/09/2021	Water	M2999-01	pH	Cool 4 deg C	PSEG03	L21	FRAC-TANK-N43502	07/09/2021	9040C

Date/Time 07/09/21 13:12
Raw Sample Received by: JP (w/c)
Raw Sample Relinquished by: JP (w/c)

Date/Time 07/09/21 14:00
Raw Sample Received by: JP (w/c)
Raw Sample Relinquished by: JP (w/c)