



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/12/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:48
In Date: 07/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:12
Out Date: 07/10/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115280

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M2864-07	VNJ-206	1	1.15	9.74	8.42	84.6	
M2864-08	12002	2	1.13	9.93	8.03	78.4	
M2936-04	TR-05-070221	3	1.17	9.74	8.69	87.7	
M2993-01	P001-B46C02-01	4	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2993-02	P001-TS12C03-01	5	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2993-03	P001-B47C04-01	6	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2993-04	P001-B47C05-01	7	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2994-01	01-A-01-B-01-C	8	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M2994-02	02-A-02-B-02-C	9	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M2997-01	RBR-200029	15	1.14	9.71	9.36	95.9	
M2997-02	RBR-200029-EPH-2	16	1.15	9.94	9.4	93.9	
M2998-01	VNJ-231	17	1.11	9.6	9.36	97.2	
M2998-02	VNJ-231-EPH-2	18	1.11	9.89	9.41	94.5	
M3000-01	AUD-1142	25	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3000-02	AUD-1143	26	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3000-03	AUD-1144	27	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3001-01	CHRT-25720	19	1.19	9.9	8.96	89.2	
M3001-02	CHRT-25720-EPH-2	20	1.13	9.91	8.91	88.6	
M3001-03	N9994	21	1.14	9.88	9.47	95.3	
M3001-04	N9994-EPH-2	22	1.12	9.62	9.44	97.9	
M3002-01	TP-F	10	1.12	9.62	8.52	87.1	
M3002-03	TP-F1	11	1.15	9.98	8.56	83.9	
M3002-04	TP-H	12	1.13	9.61	8.56	87.6	
M3002-04D	TP-HDUP	13	1.15	9.63	8.58	87.6	
M3002-06	TP-H1	14	1.16	9.97	8.94	88.3	
M3003-01	BUR-1069	28	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3005-01	BUR-1070	29	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3005-02	BUR-1071	30	1.00	2.00	2.00	100.0	WIPE SAMPLE



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/12/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:48
In Date: 07/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:12
Out Date: 07/10/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115280

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3005-03	BUR-1072	31	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3005-04	BUR-1073	32	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3006-01	CL-02-070921	23	1.18	9.65	8.56	87.1	
M3006-02	CL-02-070921-EPH-2	24	1.17	9.6	8.6	88.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$