


LB115284

7471B

INSTRUMENT ID: CV2

 μ Abs.:

52463



Linear

A= 0.0000e+000

B= 1.9093e-004

C= -8.1378e-002

Rho= 0.9997142

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.014	0.014	501	0.000	501					-
0.2	0.200	0.173	-0.027	1334	0.0 %	1334					-14
2.5	2.500	2.404	-0.096	13020	0.0 %	13020					-4
5.0	5.000	5.180	0.180	27557	0.0 %	27557					4
7.5	7.500	7.493	-0.007	39670	0.0 %	39670					0
10.0	10.000	9.935	-0.065	52463	0.0 %	52463					-1

LB115284 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Jul 2021 14:43:00

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	501	-	0.0000	7471B	PPB	09 Jul 2021 14:44:49	S	Std
0.2 - 1	50.2	1334	-	0.2000	7471B	PPB	09 Jul 2021 14:47:06	S	Std
2.5 - 1	52.5	13020	-	2.5000	7471B	PPB	09 Jul 2021 14:49:23	S	Std
5.0 - 1	55	27557	-	5.0000	7471B	PPB	09 Jul 2021 14:51:40	S	Std
7.5 - 1	57.5	39670	-	7.5000	7471B	PPB	09 Jul 2021 14:56:32	S	Std
10.0 - 1	510	52463	-	10.0000	7471B	PPB	09 Jul 2021 15:06:30	S	Std
ICV97 - 1	ICV97	21017	3.9313	-	7471B	PPB	09 Jul 2021 15:09:19	U	SMPL
ICB97 - 1	ICB97	261	-0.0315	-	7471B	PPB	09 Jul 2021 15:11:35	U	SMPL
CCV40 - 1	CCV40	27490	5.1672	-	7471B	PPB	09 Jul 2021 15:13:50	U	SMPL
CCB40 - 1	CCB40	279	-0.0281	-	7471B	PPB	09 Jul 2021 15:16:05	U	SMPL
CRA - 1	CRA	1311	0.1689	-	7471B	PPB	09 Jul 2021 15:18:21	U	SMPL
HighStd - 1	HighStd	53272	10.0897	-	7471B	PPB	09 Jul 2021 15:20:36	U	SMPL
ChkStd - 1	ChkStd	37555	7.0889	-	7471B	PPB	09 Jul 2021 15:22:52	U	SMPL
PB137606BL - 1	PBS	192	-0.0447	-	7471B	PPB	09 Jul 2021 15:25:09	U	SMPL
PB137606BS - 1	LCSS	19448	3.6318	-	7471B	PPB	09 Jul 2021 15:27:26	U	SMPL
M2940-01 - 1	LOD-MDL-SOIL-01-QT3-2021	637	0.0402	-	7471B	PPB	09 Jul 2021 15:34:45	U	SMPL
M2940-02 - 1	LOQ-SOIL-02-QT3-2021	1317	0.1701	-	7471B	PPB	09 Jul 2021 15:37:01	U	SMPL
M2989-01 - 1	OR-03-070821	2748	0.4433	-	7471B	PPB	09 Jul 2021 15:39:18	U	SMPL
M2990-01 - 1	TP-D	1398	0.1855	-	7471B	PPB	09 Jul 2021 15:41:34	U	SMPL
M2990-04 - 1	TP-G	992	0.1080	-	7471B	PPB	09 Jul 2021 15:43:50	U	SMPL
CCV41 - 1	CCV41	27339	5.1384	-	7471B	PPB	09 Jul 2021 15:46:05	U	SMPL
CCB41 - 1	CCB41	270	-0.0298	-	7471B	PPB	09 Jul 2021 15:48:21	U	SMPL
M2990-07 - 1	TP-K	1020	0.1134	-	7471B	PPB	09 Jul 2021 15:50:37	U	SMPL
M2990-10 - 1	TP-I	1015	0.1124	-	7471B	PPB	09 Jul 2021 15:52:52	U	SMPL
M2990-13 - 1	TP-J	538	0.0213	-	7471B	PPB	09 Jul 2021 15:55:08	U	SMPL
M2990-16 - 1	TP-E	2209	0.3404	-	7471B	PPB	09 Jul 2021 15:57:23	U	SMPL
M2990-16DUP - 1	TP-EDUP	2059	0.3117	-	7471B	PPB	09 Jul 2021 15:59:39	U	SMPL
M2990-16MS - 1	TP-EMS	21497	4.0230	-	7471B	PPB	09 Jul 2021 16:01:55	U	SMPL
M2990-16MSD - 1	TP-EMSD	21119	3.9508	-	7471B	PPB	09 Jul 2021 16:04:11	U	SMPL
M2990-16LX5 - 1		707	0.0536	-	7471B	PPB	09 Jul 2021 16:06:28	U	SMPL
M2990-16A - 1		23748	4.4528	-	7471B	PPB	09 Jul 2021 16:08:45	U	SMPL
CCV42 - 1	CCV42	27895	5.2445	-	7471B	PPB	09 Jul 2021 16:11:00	U	SMPL
CCB42 - 1	CCB42	283	-0.0273	-	7471B	PPB	09 Jul 2021 16:13:15	U	SMPL