

Analytical Summary ReportAnalysis Method: SM5220 DANALYST: apatelParameter: CODSUPERVISOR REVIEW BY: ketankumarRun Number: LB115289

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP90886
COD calibration std. 100 ppm	WP90884
COD calibration std. 10 ppm	WP90882
COD calibration std. 150 ppm	WP90885
COD calibration std. 50 ppm	WP90883
COD calibration std. 0 ppm	WP90881
COD CCV std, 50ppm	WP90888
COD Digestion Vials Low Level 0-150Mg/L	W2782
COD Digestion Vials Low Level 0-150Mg/L	W2824
COD ICV-LCS std, 50ppm	WP91934
COD CCV std, 50ppm	WP91933
COD LOD std, 5ppm	WP91945
COD LOQ std, 10.0PPM	WP91946

Temp In (C): <u>150</u>	Date In: <u>07/09/2021</u>	Time In: <u>13:20</u>
Temp Out (C): <u>150</u>	Date Out: <u>07/09/2021</u>	Time Out: <u>15:22</u>

Intercept: 0.2708Slope: 1.015Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.267		05/08/2021	12:05
2	CAL2	10	1	Water	11.000	10.571	-5.7	05/08/2021	12:05
3	CAL3	50	1	Water	51.000	49.98	0	05/08/2021	12:06
4	CAL4	100	1	Water	101.000	99.241	0.8	05/08/2021	12:06
5	CAL5	150	1	Water	153.000	150.472	-0.3	05/08/2021	12:07

Analysis Method: SM5220 DANALYST: apatelParameter: CODSUPERVISOR REVIEW BY: ketankumarRun Number: LB115289

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	52.000	50.965	05/08/2021	12:08
2	ICB		NA	NA	1	Water	0.000	-0.267	05/08/2021	12:08
3	CCV1	50	NA	NA	1	Water	50.000	48.994	07/09/2021	15:40
4	CCB1		NA	NA	1	Water	1.000	0.718	07/09/2021	15:41
5	LB115289BL		1.00	50	1	Solid	1.000	0.718	07/09/2021	15:42
6	LB115289BS	50	1.00	50	1	Solid	49.000	48.009	07/09/2021	15:42
7	M2940-01		1.00	50	1	Solid	5.000	4.659	07/09/2021	15:43
8	M2940-02		1.00	50	1	Solid	10.000	9.585	07/09/2021	15:43
9	CCV2	50	NA	NA	1	Water	51.000	49.980	07/09/2021	15:44
10	CCB2		NA	NA	1	Water	0.000	-0.267	07/09/2021	15:44