


LB115298

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.6177e-004

C= -9.6044e-002

Rho= 0.9999148

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/01)
0.0	0.000	0.011	0.011	410	0.000	410					-
0.2	0.200	0.181	-0.019	1060	0.0 %	1060					-10
2.5	2.500	2.456	-0.044	9751	0.0 %	9751					-2
5.0	5.000	5.100	0.100	19851	0.0 %	19851					2
7.5	7.500	7.463	-0.037	28879	0.0 %	28879					0
10.0	10.000	9.987	-0.013	38520	0.0 %	38520					0

LB115299 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jul 2021 11:42:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	410	-	0.0000	7470A	PPB	12 Jul 2021 11:59:29	S	Std
0.2 - 1	SP2	1060	-	0.2000	7470A	PPB	12 Jul 2021 12:01:46	S	Std
2.5 - 1	SP2-5	9751	-	2.5000	7470A	PPB	12 Jul 2021 12:04:03	S	Std
5.0 - 1	SP5	19851	-	5.0000	7470A	PPB	12 Jul 2021 12:06:20	S	Std
7.5 - 1	SP7-5	28879	-	7.5000	7470A	PPB	12 Jul 2021 12:08:36	S	Std
10.0 - 1	SP10	38520	-	10.0000	7470A	PPB	12 Jul 2021 12:10:52	S	Std
ICV98 - 1	ICV98	14437	3.6831	-	7470A	PPB	12 Jul 2021 12:13:49	U	SMPL
ICB98 - 1	ICB98	482	0.0301	-	7470A	PPB	12 Jul 2021 12:16:04	U	SMPL
CCV43 - 1	CCV43	17934	4.5984	-	7470A	PPB	12 Jul 2021 12:18:19	U	SMPL
CCB43 - 1	CCB43	493	0.0330	-	7470A	PPB	12 Jul 2021 12:20:34	U	SMPL
CRA - 1	CRA	1043	0.1770	-	7470A	PPB	12 Jul 2021 12:22:48	U	SMPL
HighStd - 1	HighStd	37461	9.7099	-	7470A	PPB	12 Jul 2021 12:25:04	U	SMPL
ChkStd - 1	ChkStd	26302	6.7889	-	7470A	PPB	12 Jul 2021 12:27:18	U	SMPL
PB137605BL - 1	PBW	371	0.0011	-	7470A	PPB	12 Jul 2021 12:32:13	U	SMPL
PB137605BS - 1	LCSW	14502	3.7001	-	7470A	PPB	12 Jul 2021 12:34:29	U	SMPL
M2969-02 - 1	SB01	574	0.0542	-	7470A	PPB	12 Jul 2021 12:36:45	U	SMPL
M2969-04 - 1	SB02	404	0.0097	-	7470A	PPB	12 Jul 2021 12:39:01	U	SMPL
M2969-06 - 1	SB03	210	-0.0411	-	7470A	PPB	12 Jul 2021 12:41:17	U	SMPL
M2969-08 - 1	SB04	93	-0.0717	-	7470A	PPB	12 Jul 2021 12:43:34	U	SMPL
M2969-10 - 1	SB05	716	0.0914	-	7470A	PPB	12 Jul 2021 12:45:49	U	SMPL
CCV44 - 1	CCV44	19080	4.8984	-	7470A	PPB	12 Jul 2021 12:48:05	U	SMPL
CCB44 - 1	CCB44	496	0.0338	-	7470A	PPB	12 Jul 2021 12:50:20	U	SMPL
M2969-12 - 1	SB06	653	0.0749	-	7470A	PPB	12 Jul 2021 12:52:35	U	SMPL
M2969-14 - 1	SB07	630	0.0689	-	7470A	PPB	12 Jul 2021 12:54:50	U	SMPL
M2969-16 - 1	SB08	687	0.0838	-	7470A	PPB	12 Jul 2021 12:57:06	U	SMPL
M2990-02 - 1	TP-D	403	0.0094	-	7470A	PPB	12 Jul 2021 13:06:39	U	SMPL
M2990-05 - 1	TP-G	458	0.0238	-	7470A	PPB	12 Jul 2021 13:08:55	U	SMPL
M2990-08 - 1	TP-K	286	-0.0212	-	7470A	PPB	12 Jul 2021 13:11:11	U	SMPL
M2990-11 - 1	TP-I	285	-0.0214	-	7470A	PPB	12 Jul 2021 13:13:28	U	SMPL
M2990-14 - 1	TP-J	270	-0.0254	-	7470A	PPB	12 Jul 2021 13:15:44	U	SMPL
M2990-17 - 1	TP-E	254	-0.0296	-	7470A	PPB	12 Jul 2021 13:18:01	U	SMPL
M2990-17DUP - 1	TP-EDUP	212	-0.0405	-	7470A	PPB	12 Jul 2021 13:20:17	U	SMPL
CCV45 - 1	CCV45	19155	4.9181	-	7470A	PPB	12 Jul 2021 13:22:32	U	SMPL
CCB45 - 1	CCB45	509	0.0372	-	7470A	PPB	12 Jul 2021 13:24:48	U	SMPL
M2990-17MS - 1	TP-EMS	13646	3.4760	-	7470A	PPB	12 Jul 2021 13:27:03	U	SMPL
M2990-17MSD - 1	TP-EMSD	14014	3.5723	-	7470A	PPB	12 Jul 2021 13:29:18	U	SMPL
PB137605TB - 1	PB137582TB	479	0.0293	-	7470A	PPB	12 Jul 2021 13:31:34	U	SMPL
M2990-17LX5 - 1		350	-0.0044	-	7470A	PPB	12 Jul 2021 13:33:49	U	SMPL
M2990-17A - 1		15557	3.9762	-	7470A	PPB	12 Jul 2021 13:36:05	U	SMPL
CCV46 - 1	CCV46	18529	4.7542	-	7470A	PPB	12 Jul 2021 13:38:21	U	SMPL
CCB46 - 1	CCB46	515	0.0388	-	7470A	PPB	12 Jul 2021 13:40:37	U	SMPL