



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115301

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	07/12/2021	12:45
2	CAL2	1	Water	NA	NA	21.4	7.00	07/12/2021	12:46
3	CAL3	1	Water	NA	NA	21.4	10.04	07/12/2021	12:47
4	ICV	1	Water	NA	NA	21.5	7.01	07/12/2021	12:48
5	CCV1	1	Water	NA	NA	21.4	2.01	07/12/2021	12:49
6	M3015-01	1	Water	NA	NA	25.4	7.32	07/12/2021	12:50
7	M3015-01DUP	1	Water	NA	NA	25.5	7.33	07/12/2021	12:51
8	CCV2	1	Water	NA	NA	21.4	12.02	07/12/2021	13:00

18 11530)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3015 W

WorkList ID : 150487

Department : Wet-Chemistry

Date : 07-12-2021 12:36:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/12/2021	Water	M3015-01	pH	Cool 4 deg C	WALS01	L22	RED-FRAC-TABK	07/12/2021	9040C

Date/Time 07/12/21 12:40
Raw Sample Received by: JO (WC)
Raw Sample Relinquished by: JO (WC)

Date/Time 07/12/21 13:30
Raw Sample Received by: JO (WC)
Raw Sample Relinquished by: JO (WC)