


LB115307

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.6615e-004

C= -3.7583e-002

Rho= 0.9996939

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	-0.015	-0.015	85	0.000	85					-
0.2	0.200	0.186	-0.014	841	0.0 %	841					-7
2.5	2.500	2.493	-0.007	9508	0.0 %	9508					0
5.0	5.000	5.155	0.155	19510	0.0 %	19510					.3
7.5	7.500	7.344	-0.156	27736	0.0 %	27736					-2
10.0	10.000	10.036	0.036	37849	0.0 %	37849					0

LB115307 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 12 Jul 2021 14:28:42				Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units					
0.0 - 1	50	85	-	0.0007470A	PPB			12 Jul 2021 14:30:57	S	Std
0.2 - 1	60.2	841	-	0.20007470A	PPB			12 Jul 2021 14:33:13	S	Std
2.5 - 1	52.5	9508	-	2.50007470A	PPB			12 Jul 2021 14:35:31	S	Std
5.0 - 1	55	19510	-	5.00007470A	PPB			12 Jul 2021 14:37:47	S	Std
7.5 - 1	59.5	27736	-	7.50007470A	PPB			12 Jul 2021 14:40:04	S	Std
10.0 - 1	510	37849	-	10.00007470A	PPB			12 Jul 2021 14:42:19	S	Std
ICV99 - 1	ICV99	14210	3.7445	-7470A	PPB			12 Jul 2021 14:45:46	U	SMPL
ICB99 - 1	ICB99	178	0.0098	-7470A	PPB			12 Jul 2021 14:50:34	U	SMPL
CCV47 - 1	CCV47	19557	5.1676	-7470A	PPB			12 Jul 2021 14:52:48	U	SMPL
CCB47 - 1	CCB47	184	0.0114	-7470A	PPB			12 Jul 2021 14:55:03	U	SMPL
CRA - 1	CRA	787	0.1719	-7470A	PPB			12 Jul 2021 14:57:18	U	SMPL
HighStd - 1	HighStd	37134	9.8458	-7470A	PPB			12 Jul 2021 14:59:33	U	SMPL
ChkStd - 1	ChkStd	26042	6.8936	-7470A	PPB			12 Jul 2021 15:01:47	U	SMPL
PB137633BL - 1	PBW	147	0.0015	-7470A	PPB			12 Jul 2021 15:06:37	U	SMPL
PB137633BS - 1	LCSW	14146	3.7274	-7470A	PPB			12 Jul 2021 15:08:52	U	SMPL
M2940-07 - 1	LOD-MDL-WATER-01-QT3-2021	276	0.0359	-7470A	PPB			12 Jul 2021 15:15:10	U	SMPL
M2940-08 - 1	LOQ-WATER-02-QT3-2021	965	0.2193	-7470A	PPB			12 Jul 2021 15:17:26	U	SMPL
M2999-01 - 1	FRAC-TANK-N43502	44	-0.0259	-7470A	PPB			12 Jul 2021 15:19:43	U	SMPL
M2999-01DUP - 1	FRAC-TANK-N43502DUP	62	-0.0211	-7470A	PPB			12 Jul 2021 15:22:00	U	SMPL
M2999-01MS - 1	FRAC-TANK-N43502MS	15059	3.9704	-7470A	PPB			12 Jul 2021 15:24:15	U	SMPL
CCV48 - 1	CCV48	19518	5.1572	-7470A	PPB			12 Jul 2021 15:26:30	U	SMPL
CCB48 - 1	CCB48	64	-0.0205	-7470A	PPB			12 Jul 2021 15:31:29	U	SMPL
M2999-01MSD - 1	FRAC-TANK-N43502MSD	15159	3.9971	-7470A	PPB			12 Jul 2021 15:36:25	U	SMPL
M2999-01LX5 - 1		-70	-0.0562	-7470A	PPB			12 Jul 2021 15:38:40	U	SMPL
M2999-01A - 1		15996	4.2198	-7470A	PPB			12 Jul 2021 15:40:56	U	SMPL
CCV49 - 1	CCV49	19423	5.1319	-7470A	PPB			12 Jul 2021 15:43:11	U	SMPL
CCB49 - 1	CCB49	144	0.0007	-7470A	PPB			12 Jul 2021 15:45:27	U	SMPL