



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115310

Slope : 99.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/12/2021	16:00
2	CAL2	1	Water	NA	NA	21.3	7.01	07/12/2021	16:01
3	CAL3	1	Water	NA	NA	21.3	10.03	07/12/2021	16:02
4	ICV	1	Water	NA	NA	21.4	7.00	07/12/2021	16:05
5	CCV1	1	Water	NA	NA	21.3	2.01	07/12/2021	16:09
6	M3002-01	1	Solid	20.02	20	23.8	6.98	07/12/2021	16:10
7	M3002-04	1	Solid	20.03	20	23.5	8.01	07/12/2021	16:11
8	M3012-02	1	Solid	20.04	20	24.2	8.99	07/12/2021	16:15
9	M3012-05	1	Solid	20.02	20	24.2	9.03	07/12/2021	16:16
10	M3013-02	1	Solid	20.03	20	24.0	8.60	07/12/2021	16:19
11	M3013-02DUP	1	Solid	20.04	20	24.1	8.61	07/12/2021	16:20
12	CCV2	1	Water	NA	NA	21.4	12.02	07/12/2021	16:22

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3013 WorkList ID : 150492 Department : Wet-Chemistry Date : 07-12-2021 15:47:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/09/2021	Solid	M3002-01	Corrosivity	Cool 4 deg C	PSEG03	L21	TP-F	07/09/2021	9045D
07/09/2021	Solid	M3002-04	Corrosivity	Cool 4 deg C	PSEG03	L21	TP-H	07/09/2021	9045D
07/12/2021	Solid	M3012-02	Corrosivity	Cool 4 deg C	PSEG03	L22	PS-TP1	07/12/2021	9045D
07/12/2021	Solid	M3012-05	Corrosivity	Cool 4 deg C	PSEG03	L22	PS-TP2	07/12/2021	9045D
07/12/2021	Solid	M3013-02	Corrosivity	Cool 4 deg C	PSEG03	L22	72-11995	07/12/2021	9045D

Date/Time 07/12/21 15:50
 Raw Sample Received by: JG WCE
 Raw Sample Relinquished by: JG WCE

Date/Time 07/12/21 17:20
 Raw Sample Received by: JG WCE
 Raw Sample Relinquished by: JG WCE