CHEMTECH

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB115311

Analyst By : jignesh
Supervisor Review By : apatel
Slope : 99.4
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/12/2021	16:00
2	CAL2	1	Water	NA	NA	21.3	7.01	07/12/2021	16:01
3	CAL3	1	Water	NA	NA	21.3	10.03	07/12/2021	16:02
4	ICV	1	Water	NA	NA	21.4	7.00	07/12/2021	16:05
5	CCV1	1	Water	NA	NA	21.3	2.01	07/12/2021	16:09
6	M3018-01	1	Water	NA	NA	25.7	5.57	07/12/2021	16:25
7	M3018-01DUP	1	Water	NA	NA	25.8	5.58	07/12/2021	16:26
8	CCV2	1	Water	NA	NA	21.4	12.02	07/12/2021	16:27

LB115311

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3018 **WorkList ID :** 150493 **Department :** Wet-Chemistry **Date :** 07-12-2021 15:54:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/02/2021	Water	M3018-01	pH	Cool 4 deg C	STER11	L11	POOL-WATER	07/02/2021	9040C

Date/Time 07/12/21 16:00
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JD CSM

Date/Time 07/12/21 17:00
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JP (WC)