


7470A

LB115329
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39394



A= 0.0000e+000

B= 2.5306e-004

C= -3.5905e-004

Rho= 0.9999738

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.007	0.007	28	0.000	28					
0.2	0.200	0.200	-0.000	791	0.0 %	791					
2.5	2.500	2.464	-0.036	9737	0.0 %	9737					
5.0	5.000	5.023	0.023	19852	0.0 %	19852					
7.5	7.500	7.537	0.037	29786	0.0 %	29786					
10.0	10.000	9.969	-0.031	39394	0.0 %	39394					

9/01)

1000

LB115329 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 13 Jul 2021 13:27:43					
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	28	-	0.0000	7470A	PPB	13 Jul 2021	13:43:08	S Std
0.2 - 1	50.2	791	-	0.2000	7470A	PPB	13 Jul 2021	13:45:25	S Std
2.5 - 1	52.5	9737	-	2.5000	7470A	PPB	13 Jul 2021	13:47:42	S Std
5.0 - 1	55	19852	-	5.0000	7470A	PPB	13 Jul 2021	13:49:59	S Std
7.5 - 1	57.5	29786	-	7.5000	7470A	PPB	13 Jul 2021	13:52:15	S Std
10.0 - 1	57.6	39394	-	10.0000	7470A	PPB	13 Jul 2021	13:54:31	S Std
ICV02 - 1	ICV02	15950	4.0360	-	7470A	PPB	13 Jul 2021	13:57:13	U SMPL
ICB02 - 1	ICB02	-97	-0.0249	-	7470A	PPB	13 Jul 2021	13:59:27	U SMPL
CCV53 - 1	CCV53	19640	4.9698	-	7470A	PPB	13 Jul 2021	14:01:42	U SMPL
CCB53 - 1	CCB53	-109	-0.0279	-	7470A	PPB	13 Jul 2021	14:03:57	U SMPL
CRA - 1	CRA	771	0.1948	-	7470A	PPB	13 Jul 2021	14:06:12	U SMPL
HighStd - 1	HighStd	39148	9.9066	-	7470A	PPB	13 Jul 2021	14:08:27	U SMPL
ChkStd - 1	ChkStd	27697	7.0088	-	7470A	PPB	13 Jul 2021	14:10:42	U SMPL
PB137659BL - 1	PBW	-135	-0.0345	-	7470A	PPB	13 Jul 2021	14:12:59	U SMPL
PB137659BS - 1	LCSW	15922	4.0289	-	7470A	PPB	13 Jul 2021	14:17:54	U SMPL
M3002-02 - 1	TP-F	-80	-0.0206	-	7470A	PPB	13 Jul 2021	14:20:09	U SMPL
M3002-05 - 1	TP-H	-14	-0.0039	-	7470A	PPB	13 Jul 2021	14:22:26	U SMPL
M3012-02 - 1	PS-TP1	19	0.0044	-	7470A	PPB	13 Jul 2021	14:24:43	U SMPL
M3012-05 - 1	PS-TP2	67	0.0166	-	7470A	PPB	13 Jul 2021	14:26:59	U SMPL
M3013-02 - 1	72-11995	38	0.0093	-	7470A	PPB	13 Jul 2021	14:29:14	U SMPL
CCV54 - 1	CCV54	19723	4.9908	-	7470A	PPB	13 Jul 2021	14:31:30	U SMPL
CCB54 - 1	CCB54	-107	-0.0274	-	7470A	PPB	13 Jul 2021	14:33:45	U SMPL
M3013-02DUP - 1	72-11995DUP	55	0.0136	-	7470A	PPB	13 Jul 2021	14:35:59	U SMPL
M3013-02MS - 1	72-11995MS	15311	3.8743	-	7470A	PPB	13 Jul 2021	14:38:15	U SMPL
M3013-02MSD - 1	72-11995MSD	15229	3.8536	-	7470A	PPB	13 Jul 2021	14:40:30	U SMPL
PB137660BL - 1	PBW	-78	-0.0201	-	7470A	PPB	13 Jul 2021	14:42:46	U SMPL
PB137660BS - 1	LCSW	15706	3.9743	-	7470A	PPB	13 Jul 2021	14:45:02	U SMPL
M2940-07 - 1	LOD-MDL-WATER-01-QT3-2021	652	0.1646	-	7470A	PPB	13 Jul 2021	14:49:51	U SMPL
M2940-09 - 1	MDL-WATER-03-QT3-2021	197	0.0495	-	7470A	PPB	13 Jul 2021	14:52:08	U SMPL
M3015-01 - 1	RED-FRAC-TANK	31	0.0075	-	7470A	PPB	13 Jul 2021	14:54:25	U SMPL
M3015-01DUP - 1	RED-FRAC-TANKDUP	30	0.0072	-	7470A	PPB	13 Jul 2021	14:56:42	U SMPL
M3015-01MS - 1	RED-FRAC-TANKMS	15538	3.9318	-	7470A	PPB	13 Jul 2021	14:58:57	U SMPL
CCV55 - 1	CCV55	19279	4.8785	-	7470A	PPB	13 Jul 2021	15:01:13	U SMPL
CCB55 - 1	CCB55	-105	-0.0269	-	7470A	PPB	13 Jul 2021	15:03:29	U SMPL
M3015-01MSD - 1	RED-FRAC-TANKMSD	14553	3.6825	-	7470A	PPB	13 Jul 2021	15:05:44	U SMPL
PB137659TB - 1	PB137638TB	-65	-0.0168	-	7470A	PPB	13 Jul 2021	15:08:00	U SMPL
M3013-02LX5 - 1		74	0.0184	-	7470A	PPB	13 Jul 2021	15:10:15	U SMPL
M3013-02A - 1		15130	3.8285	-	7470A	PPB	13 Jul 2021	15:12:30	U SMPL
M3015-01LX5 - 1		-11	-0.0031	-	7470A	PPB	13 Jul 2021	15:14:46	U SMPL
M3015-01A - 1		15722	3.9783	-	7470A	PPB	13 Jul 2021	15:17:02	U SMPL
CCV56 - 1	CCV56	19609	4.9620	-	7470A	PPB	13 Jul 2021	15:22:13	U SMPL
CCB56 - 1	CCB56	-101	-0.0259	-	7470A	PPB	13 Jul 2021	15:24:29	U SMPL