


LB 113 337

7471B

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 1.8639e-004

C= -3.4238e-002

Rho= 0.9999547

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.023	-0.023	60	0.000	60					-16
0.2	0.200	0.168	-0.032	1087	0.0 %	1087					1
2.5	2.500	2.535	0.035	13784	0.0 %	13784					1
5.0	5.000	5.051	0.051	27285	0.0 %	27285					0
7.5	7.500	7.512	0.012	40487	0.0 %	40487					0
10.0	10.000	9.957	-0.043	53603	0.0 %	53603					0

LB115337 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jul 2021 15:51:52

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50-2	60	-	0.00007471B	PPB	13 Jul 2021 15:53:17	S	Std
0.2 - 1	52-5	1087	-	0.20007471B	PPB	13 Jul 2021 15:55:34	S	Std
2.5 - 1	55	13784	-	2.50007471B	PPB	13 Jul 2021 15:57:50	S	Std
5.0 - 1	57-5	27285	-	5.00007471B	PPB	13 Jul 2021 16:00:07	S	Std
7.5 - 1	510	40487	-	7.50007471B	PPB	13 Jul 2021 16:02:23	S	Std
10.0 - 1		53603	-	10.00007471B	PPB	13 Jul 2021 16:04:39	S	Std
ICV03 - 1	ICV03	22148	4.0938	-7471B	PPB	13 Jul 2021 16:07:39	U	SMPL
ICB03 - 1	ICB03	-73	-0.0478	-7471B	PPB	13 Jul 2021 16:09:54	U	SMPL
CCV57 - 1	CCV57	27170	5.0298	-7471B	PPB	13 Jul 2021 16:12:10	U	SMPL
CCB57 - 1	CCB57	45	-0.0259	-7471B	PPB	13 Jul 2021 16:16:59	U	SMPL
CRA - 1	CRA	1222	0.1935	-7471B	PPB	13 Jul 2021 16:21:43	U	SMPL
HighStd - 1	HighStd	53612	9.9582	-7471B	PPB	13 Jul 2021 16:23:58	U	SMPL
ChkStd - 1	ChkStd	37805	7.0121	-7471B	PPB	13 Jul 2021 16:26:13	U	SMPL
PB137661BL - 1	PBS	167	-0.0031	-7471B	PPB	13 Jul 2021 16:31:02	U	SMPL
PB137661BS - 1	LCSS	22388	4.1386	-7471B	PPB	13 Jul 2021 16:33:17	U	SMPL
M2940-01 - 1	LOD-MDL-SOIL-01-QT3-2021	895	0.1326	-7471B	PPB	13 Jul 2021 16:55:47	U	SMPL
M2940-03 - 1	MDL-SOIL-03-QT3-2021	454	0.0504	-7471B	PPB	13 Jul 2021 17:03:08	U	SMPL
M3012-01 - 1	PS-TP1	4145	0.7383	-7471B	PPB	13 Jul 2021 17:05:24	U	SMPL
M3012-01DUP - 1	PS-TP1DUP	4121	0.7339	-7471B	PPB	13 Jul 2021 17:07:41	U	SMPL
M3012-01MS - 1	PS-TP1MS	25535	4.7251	-7471B	PPB	13 Jul 2021 17:09:56	U	SMPL
CCV58 - 1	CCV58	27264	5.0474	-7471B	PPB	13 Jul 2021 17:12:11	U	SMPL
CCB58 - 1	CCB58	-33	-0.0404	-7471B	PPB	13 Jul 2021 17:14:26	U	SMPL
M3012-01MSD - 1	PS-TP1MSD	25308	4.6828	-7471B	PPB	13 Jul 2021 17:16:42	U	SMPL
M3012-04 - 1	PS-TP2	54468	10.1178	-7471B	PPB	13 Jul 2021 17:18:57	U	SMPL
M3013-01 - 1	72-11995	1703	0.2832	-7471B	PPB	13 Jul 2021 17:21:13	U	SMPL
M3014-01 - 1	HD-02-071221	1123	0.1751	-7471B	PPB	13 Jul 2021 17:23:31	U	SMPL
M3012-01LX5 - 1		786	0.1123	-7471B	PPB	13 Jul 2021 17:25:47	U	SMPL
M3012-01A - 1		25481	4.7150	-7471B	PPB	13 Jul 2021 17:28:02	U	SMPL
CCV59 - 1	CCV59	27248	5.0444	-7471B	PPB	13 Jul 2021 17:30:19	U	SMPL
CCB59 - 1	CCB59	10	-0.0324	-7471B	PPB	13 Jul 2021 17:34:53	U	SMPL