



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115343

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/13/2021	17:05
2	CAL2	1	Water	NA	NA	21.2	7.00	07/13/2021	17:06
3	CAL3	1	Water	NA	NA	21.2	10.03	07/13/2021	17:07
4	ICV	1	Water	NA	NA	21.3	6.99	07/13/2021	17:10
5	CCV1	1	Water	NA	NA	21.2	2.01	07/13/2021	17:11
6	M3021-01	1	Solid	20.01	20	22.8	7.13	07/14/2021	17:13
7	M3021-03	1	Solid	20.03	20	23.0	6.76	07/13/2021	17:15
8	M3026-02	1	Solid	20.02	20	23.6	12.09	07/13/2021	17:19
9	M3026-05	1	Solid	20.03	20	23.8	12.17	07/13/2021	17:20
10	M3026-08	1	Solid	20.01	20	23.8	12.03	07/13/2021	17:22
11	M3026-11	1	Solid	20.03	20	24.0	12.20	07/13/2021	17:24
12	M3026-11DUP	1	Solid	20.04	20	24.1	12.21	07/14/2021	17:25
13	M3027-01	1	Solid	20.02	20	23.0	7.38	07/13/2021	17:26
14	CCV2	1	Water	NA	NA	21.5	12.03	07/13/2021	17:29

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3026 S WorkList ID : 150530 Department : Wet-Chemistry Date : 07-13-2021 14:58:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/13/2021	Solid	M3021-01	pH	Cool 4 deg C	PSEG03	L23	31455	07/13/2021	9045D
07/13/2021	Solid	M3021-03	pH	Cool 4 deg C	PSEG03	L23	31456	07/13/2021	9045D
07/13/2021	Solid	M3026-02	pH	Cool 4 deg C	JPCL01	L11	ABN-JG-1	07/13/2021	9045D
07/13/2021	Solid	M3026-05	pH	Cool 4 deg C	JPCL01	L11	ABN-JG-2	07/13/2021	9045D
07/13/2021	Solid	M3026-08	pH	Cool 4 deg C	JPCL01	L11	ABN-JG-3	07/13/2021	9045D
07/13/2021	Solid	M3026-11	pH	Cool 4 deg C	JPCL01	L11	ABN-JG-4	07/13/2021	9045D
07/13/2021	Solid	M3027-01	pH	Cool 4 deg C	PSEG03	L23	NWB-1572	07/13/2021	9045D

Date/Time 07-13-21 16:10
 Raw Sample Received by: JP 1202
 Raw Sample Relinquished by: CR SM

Date/Time 07-13-21 17:40
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)