



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115345

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/13/2021	17:05
2	CAL2	1	Water	NA	NA	21.2	7.00	07/13/2021	17:06
3	CAL3	1	Water	NA	NA	21.2	10.03	07/13/2021	17:07
4	ICV	1	Water	NA	NA	21.3	6.99	07/13/2021	17:10
5	CCV1	1	Water	NA	NA	21.2	2.01	07/13/2021	17:11
6	M3027-03	1	Water	NA	NA	23.0	6.16	07/13/2021	17:35
7	M3027-05	1	Water	NA	NA	22.3	7.54	07/13/2021	17:36
8	M3027-05DUP	1	Water	NA	NA	22.4	7.55	07/13/2021	17:39
9	CCV2	1	Water	NA	NA	21.3	12.02	07/14/2021	17:40

WORKLIST(Hardcopy Internal Chain)

VB 115345

WorkList Name : PH M3027 W WorkList ID : 150531 Department : Wet-Chemistry Date : 07-13-2021 14:58:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/13/2021	Water	M3027-03	pH	Cool 4 deg C	PSEG03	L23	NWB-1639	07/13/2021	9040C
07/13/2021	Water	M3027-05	pH	Cool 4 deg C	PSEG03	L23	NWB-1640	07/13/2021	9040C

Date/Time 07-13-21 16:10
 Raw Sample Received by: JQ (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 07-13-21 17:40
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JQ (WC)