

LB 115356


7471B

INSTRUMENT ID: CVI

Linear

 μ Abs.:

53693



A= 0.0000e+000

B= 1.8670e-004

C= -8.4709e-003

Rho= 0.9999150

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.006	-0.006	11	0.000	11					
0.2	0.200	0.192	-0.008	1074	0.0 %	1074					
0.5	2.500	2.489	-0.011	13376	0.0 %	13376					
1.0	5.000	5.086	0.086	27289	0.0 %	27289					
1.5	7.500	7.423	-0.077	39804	0.0 %	39804					
2.0	10.000	10.016	0.016	53693	0.0 %	53693					

9013
 -4
 0
 2
 -1
 0

LB115356 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Jul 2021 16:01:37

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	11	-	0.00007471B	PPB		14 Jul 2021 16:05:38	S	Std
0.2 - 1	50.2	1074	-	0.20007471B	PPB		14 Jul 2021 16:07:55	S	Std
2.5 - 1	52.5	13376	-	2.50007471B	PPB		14 Jul 2021 16:10:12	S	Std
5.0 - 1	55	27289	-	5.00007471B	PPB		14 Jul 2021 16:12:29	S	Std
7.5 - 1	57.5	39804	-	7.50007471B	PPB		14 Jul 2021 16:14:45	S	Std
10.0 - 1	59	53693	-	10.00007471B	PPB		14 Jul 2021 16:17:02	S	Std
ICV05 - 1	ICV05	21309	3.9700	-7471B	PPB		14 Jul 2021 16:19:39	U	SMPL
ICB05 - 1	ICB05	-80	-0.0234	-7471B	PPB		14 Jul 2021 16:21:54	U	SMPL
CCV64 - 1	CCV64	27458	5.1180	-7471B	PPB		14 Jul 2021 16:24:09	U	SMPL
CCB64 - 1	CCB64	-85	-0.0243	-7471B	PPB		14 Jul 2021 16:26:24	U	SMPL
CRA - 1	CRA	1022	0.1823	-7471B	PPB		14 Jul 2021 16:28:40	U	SMPL
HighStd - 1	HighStd	53671	10.0120	-7471B	PPB		14 Jul 2021 16:30:55	U	SMPL
ChkStd - 1	ChkStd	37892	7.0660	-7471B	PPB		14 Jul 2021 16:35:46	U	SMPL
PB137693BL - 1	PBS	-151	-0.0367	-7471B	PPB		14 Jul 2021 16:38:02	U	SMPL
PB137693BS - 1	LCSS	22769	4.2425	-7471B	PPB		14 Jul 2021 16:40:19	U	SMPL
M3022-01 - 1	ETGI-363	3910	0.7215	-7471B	PPB		14 Jul 2021 16:42:35	U	SMPL
M3023-01 - 1	OR-02-071321	4901	0.9066	-7471B	PPB		14 Jul 2021 16:44:52	U	SMPL
M3029-01 - 1	RVR-CF01-01	1884	0.3433	-7471B	PPB		14 Jul 2021 16:47:09	U	SMPL
M3029-01DUP - 1	RVR-CF01-01DUP	1806	0.3287	-7471B	PPB		14 Jul 2021 16:49:26	U	SMPL
M3029-02 - 1	M3029-01MS	22492	4.1908	-7471B	PPB		14 Jul 2021 16:51:41	U	SMPL
CCV65 - 1	CCV65	27170	5.0642	-7471B	PPB		14 Jul 2021 16:53:56	U	SMPL
CCB65 - 1	CCB65	-101	-0.0273	-7471B	PPB		14 Jul 2021 16:56:11	U	SMPL
M3029-03 - 1	M3029-01MSD	23123	4.3086	-7471B	PPB		14 Jul 2021 16:58:26	U	SMPL
M3029-04 - 1	RVR-CF01-02	1640	0.2977	-7471B	PPB		14 Jul 2021 17:00:41	U	SMPL
M3029-05 - 1	RVR-CF02-01	2060	0.3761	-7471B	PPB		14 Jul 2021 17:02:57	U	SMPL
M3029-06 - 1	RVR-CF03-01	1909	0.3479	-7471B	PPB		14 Jul 2021 17:05:12	U	SMPL
M3029-07 - 1	RVR-CF04-01	1894	0.3451	-7471B	PPB		14 Jul 2021 17:07:28	U	SMPL
M3029-01LX5 - 1		353	0.0574	-7471B	PPB		14 Jul 2021 17:09:45	U	SMPL
M3029-01A - 1		23310	4.3435	-7471B	PPB		14 Jul 2021 17:12:01	U	SMPL
M3012-04DLX2 - 1		27293	5.0872	-7471B	PPB		14 Jul 2021 17:14:17	U	SMPL
CV66 - 1	CCV66	27049	5.0416	-7471B	PPB		14 Jul 2021 17:16:34	U	SMPL
CB66 - 1	CCB66	-82	-0.0238	-7471B	PPB		14 Jul 2021 17:18:49	U	SMPL