

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 07/15/2021

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 103 °C

DATE & TIME IN1: 07/14/2021 16:30

DATE & TIME OUT1: 07/15/2021 08:20

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 07/15/2021 09:00

DATE & TIME OUT2: 07/15/2021 10:30

Run Number: LB115358

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB115358BL	LB115358BL	143.5971	143.5971	100	143.5972	143.5972	143.5972	0.0001	1
2	M3039-22	EX-05	152.9791	152.9791	100	153.1093	153.1093	153.1093	0.1302	1302
3	M3039-22DUP	EX-05DUP	151.2289	151.2289	100	151.3571	151.3571	151.3571	0.1282	1282

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

VB 115358

WorkList Name : ts m3039 WorkList ID : 150567 Department : Wet-Chemistry Date : 07-14-2021 16:06:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/21/2021	Water	M3039-22	TS	Cool 4 deg C	LIRO01	L31	EX-05	07/14/2021	SM2540 B

Date/Time 07-14-21 16:10
 Raw Sample Received by: JR (WCL)
 Raw Sample Relinquished by: JD CSM

Date/Time 07-14-21 17:30
 Raw Sample Received by: JD CSM
 Raw Sample Relinquished by: JD CSM