



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115359

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/14/2021	16:20
2	CAL2	1	Water	NA	NA	21.2	7.01	07/14/2021	16:21
3	CAL3	1	Water	NA	NA	21.2	10.03	07/14/2021	16:22
4	ICV	1	Water	NA	NA	21.3	7.00	07/14/2021	16:23
5	CCV1	1	Water	NA	NA	21.3	2.01	07/14/2021	16:25
6	M3038-13	1	Solid	20.02	20	24.2	7.00	07/14/2021	16:26
7	M3038-13DUP	1	Solid	20.03	20	24.3	7.01	07/14/2021	16:27
8	M3038-14	1	Solid	20.04	20	24.3	7.84	07/14/2021	16:30
9	M3038-15	1	Solid	20.03	20	24.2	6.87	07/14/2021	16:32
10	M3038-16	1	Solid	20.02	20	24.0	7.71	07/14/2021	16:35
11	M3038-17	1	Solid	20.01	20	23.8	10.32	07/14/2021	16:36
12	M3039-15	1	Solid	20.03	20	24.0	10.77	07/14/2021	16:38
13	M3039-16	1	Solid	20.04	20	24.0	8.56	07/14/2021	16:40
14	M3039-17	1	Solid	20.02	20	23.9	8.74	07/14/2021	16:41
15	M3039-18	1	Solid	20.03	20	22.9	9.02	07/14/2021	16:42
16	CCV2	1	Water	NA	NA	21.4	12.02	07/14/2021	16:44
17	M3039-19	1	Solid	20.04	20	24.0	8.42	07/14/2021	16:45
18	M3039-20	1	Solid	20.02	20	23.9	7.82	07/14/2021	16:47
19	M3039-21	1	Solid	20.03	20	23.8	7.39	07/14/2021	16:48
20	CCV3	1	Water	NA	NA	21.4	2.01	07/14/2021	16:50

WORKLIST(Hardcopy Internal Chain)

VB 115359

WorkList Name : CORROSIONITY M3038 WorkList ID : 150569 Department : Wet-Chemistry Date : 07-14-2021 16:10:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/09/2021	Solid	M3038-13	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-06-COMP	07/09/2021	9045D
07/09/2021	Solid	M3038-14	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-07-COMP	07/09/2021	9045D
07/09/2021	Solid	M3038-15	Corrosivity	Cool 4 deg C	LIRO01	L31	SB09-COMP	07/09/2021	9045D
07/09/2021	Solid	M3038-16	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-11-COMP	07/09/2021	9045D
07/12/2021	Solid	M3038-17	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-10-COMP	07/12/2021	9045D
07/12/2021	Solid	M3039-15	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-12-COMP	07/12/2021	9045D
07/13/2021	Solid	M3039-16	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-01-COMP	07/13/2021	9045D
07/13/2021	Solid	M3039-17	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-02-COMP	07/13/2021	9045D
07/13/2021	Solid	M3039-18	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-03-COMP	07/13/2021	9045D
07/13/2021	Solid	M3039-19	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-04-COMP	07/13/2021	9045D
07/13/2021	Solid	M3039-20	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-05-COMP	07/13/2021	9045D
07/13/2021	Solid	M3039-21	Corrosivity	Cool 4 deg C	LIRO01	L31	SB-08-COMP	07/13/2021	9045D

Date/Time 07-14-21 16:15
 Raw Sample Received by: JR (WC)
 Raw Sample Relinquished by: JR (WC)

Date/Time 07-14-21 17:35
 Raw Sample Received by: JR (WC)
 Raw Sample Relinquished by: JR (WC)