



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115360

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/14/2021	16:20
2	CAL2	1	Water	NA	NA	21.2	7.01	07/14/2021	16:21
3	CAL3	1	Water	NA	NA	21.2	10.03	07/14/2021	16:22
4	ICV	1	Water	NA	NA	21.3	7.00	07/14/2021	16:23
5	CCV1	1	Water	NA	NA	21.3	2.01	07/14/2021	16:25
6	M3040-01	1	Solid	20.02	20	24.3	6.97	07/14/2021	17:21
7	M3040-01DUP	1	Solid	20.03	20	24.4	6.98	07/14/2021	17:22
8	M3043-01	1	Solid	20.04	20	24.4	7.23	07/14/2021	17:25
9	CCV2	1	Water	NA	NA	21.3	12.02	07/14/2021	17:26

WORKLIST(Hardcopy Internal Chain)

✓ 15360

WorkList Name : PH M3043

WorkList ID : 150570

Department : Wet-Chemistry

Date : 07-14-2021 17:26:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/14/2021	Solid	M3040-01	pH	Cool 4 deg C	TULL02	L31	FK-PS-01-076	07/14/2021	9045D
07/14/2021	Solid	M3043-01	pH	Cool 4 deg C	TULL02	L31	FK-PS-01-077	07/14/2021	9045D

Date/Time 07-14-21 16:15
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: SP (WC)

Date/Time 07-14-21 17:35
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: SP (WC)