



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115370

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/15/2021	14:00
2	CAL2	1	Water	NA	NA	21.3	7.00	07/15/2021	14:01
3	CAL3	1	Water	NA	NA	21.3	10.03	07/15/2021	14:02
4	ICV	1	Water	NA	NA	21.4	6.99	07/15/2021	14:05
5	CCV1	1	Water	NA	NA	21.3	2.01	07/15/2021	14:06
6	M3057-01	1	Water	NA	NA	22.5	8.15	07/15/2021	14:10
7	M3057-01DUP	1	Water	NA	NA	22.6	8.16	07/15/2021	14:11
8	M3057-02	1	Water	NA	NA	23.1	8.35	07/15/2021	14:13
9	M3057-03	1	Water	NA	NA	23.6	8.37	07/15/2021	14:15
10	M3064-01	1	Water	NA	NA	22.8	5.70	07/15/2021	14:19
11	CCV2	1	Water	NA	NA	21.4	12.02	07/15/2021	14:20

WORKLIST(Hardcopy Internal Chain)

18115310

WorkList Name : PH M3064 WorkList ID : 150588 Department : Wet-Chemistry Date : 07-15-2021 13:49:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/15/2021	Water	M3057-01	pH	Cool 4 deg C	RJIN01	L11	EQ-TANK-TESTING-TOP	07/15/2021	9040C
07/15/2021	Water	M3057-02	pH	Cool 4 deg C	RJIN01	L11	EQ-TANK-TESTING-MIDI	07/15/2021	9040C
07/15/2021	Water	M3057-03	pH	Cool 4 deg C	RJIN01	L11	EQ-TANK-TESTING-BOT	07/15/2021	9040C
07/12/2021	Water	M3064-01	pH	Cool 4 deg C	STER10	L11	POOL-WATER	07/12/2021	9040C

Date/Time 07/15/21 13:50
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CR SM

Date/Time 07/15/21 15:30
 Raw Sample Received by: CR SM
 Raw Sample Relinquished by: JP (WC)