



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115371

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	07/15/2021	14:00
2	CAL2	1	Water	NA	NA	21.3	7.00	07/15/2021	14:01
3	CAL3	1	Water	NA	NA	21.3	10.03	07/15/2021	14:02
4	ICV	1	Water	NA	NA	21.4	6.99	07/15/2021	14:05
5	CCV1	1	Water	NA	NA	21.3	2.01	07/15/2021	14:06
6	M3061-02	1	Water	NA	NA	22.4	11.33	07/15/2021	14:25
7	M3061-02DUP	1	Water	NA	NA	22.5	11.34	07/15/2021	14:26
8	CCV2	1	Water	NA	NA	21.3	12.01	07/15/2021	14:29

WORKLIST(Hardcopy Internal Chain)

✓ 115371

WorkList Name : CORROSIONITY M3061 V WorkList ID : 150591 Department : Wet-Chemistry Date : 07-15-2021 14:16:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/14/2021	Water	M3061-02	E Corrosivity	Cool 4 deg C	ARCA05	L11	PB-2021-WS-000-08	07/14/2021	9040C

Date/Time 07/15/21 14:20
 Raw Sample Received by: JO (WC)
 Raw Sample Relinquished by: CR SM

Date/Time 07/15/21 15:30
 Raw Sample Received by: CR SM
 Raw Sample Relinquished by: JO (WC)