

LB115372


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53769



A= 0.0000e+000

B= 1.8457e-004

C= -1.7447e-002

Rho= 0.9998648

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
1.0	0.000	-0.004	-0.004	73	0.000	73					
1.2	0.200	0.177	-0.023	1053	0.0 %	1053					-12
1.5	2.500	2.479	-0.021	13524	0.0 %	13524					-1
1.0	5.000	5.037	0.037	27388	0.0 %	27388					-1
1.5	7.500	7.605	0.105	41297	0.0 %	41297					-1
0.0	10.000	9.906	-0.094	53769	0.0 %	53769					-1

LB115372 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Jul 2021 14:55:35

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	73	-	0.0000	7471B	PPB	15 Jul 2021 14:57:17	S	Std
0.2 - 1	50-2	1053	-	0.2000	7471B	PPB	15 Jul 2021 14:59:34	S	Std
2.5 - 1	52-5	13524	-	2.5000	7471B	PPB	15 Jul 2021 15:01:51	S	Std
5.0 - 1	55	27388	-	5.0000	7471B	PPB	15 Jul 2021 15:04:08	S	Std
7.5 - 1	57-5	41297	-	7.5000	7471B	PPB	15 Jul 2021 15:06:24	S	Std
10.0 - 1	57-5	53769	-	10.0000	7471B	PPB	15 Jul 2021 15:08:40	S	Std
ICV07 - 1	ICV07	22603	4.1543	-	7471B	PPB	15 Jul 2021 15:11:31	U	SMPL
ICB07 - 1	ICB07	-74	-0.0311	-	7471B	PPB	15 Jul 2021 15:13:46	U	SMPL
CCV70 - 1	CCV70	27440	5.0470	-	7471B	PPB	15 Jul 2021 15:16:01	U	SMPL
CCB70 - 1	CCB70	-71	-0.0306	-	7471B	PPB	15 Jul 2021 15:18:16	U	SMPL
CRA - 1	CRA	1041	0.1747	-	7471B	PPB	15 Jul 2021 15:20:32	U	SMPL
HighStd - 1	HighStd	53414	9.8409	-	7471B	PPB	15 Jul 2021 15:22:47	U	SMPL
ChkStd - 1	ChkStd	38103	7.0151	-	7471B	PPB	15 Jul 2021 15:25:02	U	SMPL
PB137716BL - 1	PBS	-176	-0.0499	-	7471B	PPB	15 Jul 2021 15:27:20	U	SMPL
PB137716BS - 1	LCSS	19238	3.5332	-	7471B	PPB	15 Jul 2021 15:29:37	U	SMPL
M3026-02 - 1	ABN-JG-1	14436	2.6469	-	7471B	PPB	15 Jul 2021 15:31:53	U	SMPL
M3026-05 - 1	ABN-JG-2	37846	6.9676	-	7471B	PPB	15 Jul 2021 15:34:09	U	SMPL
M3026-08 - 1	ABN-JG-3	13208	2.4203	-	7471B	PPB	15 Jul 2021 15:36:26	U	SMPL
M3026-11 - 1	ABN-JG-4	12876	2.3590	-	7471B	PPB	15 Jul 2021 15:38:44	U	SMPL
M3031-01 - 1	HR-03-071421	5011	0.9074	-	7471B	PPB	15 Jul 2021 15:40:59	U	SMPL
CCV71 - 1	CCV71	27012	4.9680	-	7471B	PPB	15 Jul 2021 15:43:14	U	SMPL
CCB71 - 1	CCB71	-71	-0.0306	-	7471B	PPB	15 Jul 2021 15:45:30	U	SMPL
M3032-01 - 1	NB-07-071421	852	0.1398	-	7471B	PPB	15 Jul 2021 15:47:45	U	SMPL
M3037-01 - 1	DS-1	130931	24.1479	-	7471B	PPB	15 Jul 2021 15:50:00	U	SMPL
M3037-02 - 1	DN-1	63599	11.7207	-	7471B	PPB	15 Jul 2021 15:52:15	U	SMPL
M3037-03 - 1	CS-1	61890	11.4053	-	7471B	PPB	15 Jul 2021 15:54:33	U	SMPL
M3037-04 - 1	CN-1	60336	11.1185	-	7471B	PPB	15 Jul 2021 15:56:51	U	SMPL
M3037-05 - 1	BN-1	100691	18.5666	-	7471B	PPB	15 Jul 2021 15:59:08	U	SMPL
M3037-06 - 1	AN-1	84672	15.6101	-	7471B	PPB	15 Jul 2021 16:01:27	U	SMPL
M3040-01 - 1	FK-PS-01-076	1224	0.2085	-	7471B	PPB	15 Jul 2021 16:03:46	U	SMPL
M3043-01 - 1	FK-PS-01-077	2090	0.3683	-	7471B	PPB	15 Jul 2021 16:06:05	U	SMPL
M3046-01 - 1	FK-EF-01-019-ABCD	1543	0.2673	-	7471B	PPB	15 Jul 2021 16:08:20	U	SMPL
CCV72 - 1	CCV72	27501	5.0583	-	7471B	PPB	15 Jul 2021 16:10:35	U	SMPL
CCB72 - 1	CCB72	-79	-0.0320	-	7471B	PPB	15 Jul 2021 16:12:51	U	SMPL
M3047-01 - 1	FK-EF-01-020-ABCD	2522	0.4480	-	7471B	PPB	15 Jul 2021 16:15:06	U	SMPL
M3048-01 - 1	FK-BPM-01-AA1-AA1A-BB1-BB1A	955	0.1588	-	7471B	PPB	15 Jul 2021 16:17:21	U	SMPL
M3049-01 - 1	FK-BPM-03-024-ABCD	1412	0.2432	-	7471B	PPB	15 Jul 2021 16:19:37	U	SMPL
M3050-01 - 1	FK-BPM-04-018-ABCD	760	0.1228	-	7471B	PPB	15 Jul 2021 16:21:52	U	SMPL
M3050-01DUP - 1	FK-BPM-04-018-ABCD	822	0.1343	-	7471B	PPB	15 Jul 2021 16:26:52	U	SMPL
M3050-01MS - 1	FK-BPM-04-018-ABCDMS	22419	4.1203	-	7471B	PPB	15 Jul 2021 16:29:08	U	SMPL
M3050-01MSD - 1	FK-BPM-04-018-ABCDMSD	20496	3.7654	-	7471B	PPB	15 Jul 2021 16:31:24	U	SMPL
M3050-01LX5 - 1		110	0.0029	-	7471B	PPB	15 Jul 2021 16:33:40	U	SMPL
M3050-01A - 1		22345	4.1067	-	7471B	PPB	15 Jul 2021 16:35:57	U	SMPL
ICV73 - 1	CCV73	27229	5.0081	-	7471B	PPB	15 Jul 2021 16:38:12	U	SMPL
ICB73 - 1	CCB73	-85	-0.0331	-	7471B	PPB	15 Jul 2021 16:40:28	U	SMPL
I3037-01DLX5 - 1		25198	4.6332	-	7471B	PPB	15 Jul 2021 16:42:43	U	SMPL
I3037-02DLX2 - 1		34249	6.3037	-	7471B	PPB	15 Jul 2021 16:44:59	U	SMPL
I3037-03DLX2 - 1		33368	6.1411	-	7471B	PPB	15 Jul 2021 16:47:14	U	SMPL
I3037-04DLX2 - 1		30996	5.7033	-	7471B	PPB	15 Jul 2021 16:49:30	U	SMPL
I3037-05DLX5 - 1		19528	3.5867	-	7471B	PPB	15 Jul 2021 16:51:46	U	SMPL
I3037-06DLX5 - 1		16480	3.0242	-	7471B	PPB	15 Jul 2021 16:54:02	U	SMPL
CV74 - 1	CCV74	26204	4.8189	-	7471B	PPB	15 Jul 2021 16:56:18	U	SMPL
CB74 - 1	CCB74	-86	-0.0333	-	7471B	PPB	15 Jul 2021 16:58:34	U	SMPL