



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115377

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	07/16/2021	08:15
2	CAL2	1	Water	NA	NA	21.1	7.01	07/16/2021	08:16
3	CAL3	1	Water	NA	NA	21.1	10.03	07/16/2021	08:17
4	ICV	1	Water	NA	NA	21.2	6.98	07/16/2021	08:19
5	CCV1	1	Water	NA	NA	21.2	2.01	07/16/2021	08:20
6	M3051-08	1	Solid	20.02	20	24.9	7.70	07/16/2021	08:33
7	M3051-08DUP	1	Solid	20.03	20	24.8	7.71	07/16/2021	08:34
8	CCV2	1	Water	NA	NA	21.2	12.02	07/16/2021	08:35

WORKLIST(Hardcopy Internal Chain)

2B115347

WorkList Name : ph m3051 solids WorkList ID : 150595 Department : Wet-Chemistry Date : 07-16-2021 08:02:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/15/2021	Solid	M3051-08	pH	Cool 4 deg C	PSEG03	L33	41213	07/15/2021	9045D

Date/Time 07/16/21 08:10
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 07/16/21 10:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)