

LB115378


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

3743E



A= 0.0000e+000

B= 2.6233e-004

C= 1.5649e-002

Rho= 0.9995772

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
1.0	0.000	-0.002	-0.002	-68	0.000	-68					-15
1.2	0.200	0.170	-0.030	587	0.0 %	587					-2
1.5	2.500	2.442	-0.058	9250	0.0 %	9250					1
1.0	5.000	5.074	0.074	19284	0.0 %	19284					2
1.5	7.500	7.680	0.180	29216	0.0 %	29216					-2
0.0	10.000	9.836	-0.164	37435	0.0 %	37435					

LB115378 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Jul 2021 09:42:31

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-68	-	0.0000	7470A	PPB	16 Jul 2021 09:54:33	S	Std
0.2 - 1	50.2	587	-	0.2000	7470A	PPB	16 Jul 2021 09:56:50	S	Std
2.5 - 1	52.5	9250	-	2.5000	7470A	PPB	16 Jul 2021 09:59:07	S	Std
5.0 - 1	55	19284	-	5.0000	7470A	PPB	16 Jul 2021 10:01:24	S	Std
7.5 - 1	57.5	29216	-	7.5000	7470A	PPB	16 Jul 2021 10:03:40	S	Std
10.0 - 1	510	37435	-	10.0000	7470A	PPB	16 Jul 2021 10:08:58	S	Std
CV08 - 1	ICV08	14675	3.8654	-7470A	PPB	16 Jul 2021 10:11:41	U	SMPL	
CB08 - 1	ICB08	-118	-0.0153	-7470A	PPB	16 Jul 2021 10:13:55	U	SMPL	
CV75 - 1	CCV75	18829	4.9551	-7470A	PPB	16 Jul 2021 10:16:10	U	SMPL	
CB75 - 1	CCB75	-109	-0.0129	-7470A	PPB	16 Jul 2021 10:18:25	U	SMPL	
RA - 1	CRA	615	0.1770	-7470A	PPB	16 Jul 2021 10:20:40	U	SMPL	
HighStd - 1	HighStd	37634	9.8882	-7470A	PPB	16 Jul 2021 10:22:55	U	SMPL	
ChkStd - 1	ChkStd	26948	7.0849	-7470A	PPB	16 Jul 2021 10:25:10	U	SMPL	
B137717BL - 1	PBW	-160	-0.0263	-7470A	PPB	16 Jul 2021 10:27:27	U	SMPL	
B137717BS - 1	LCSW	13994	3.6867	-7470A	PPB	16 Jul 2021 10:29:43	U	SMPL	
3038-13 - 1	SB-06-COMP	-76	-0.0043	-7470A	PPB	16 Jul 2021 10:31:59	U	SMPL	
3038-14 - 1	SB-07-COMP	11	0.0185	-7470A	PPB	16 Jul 2021 10:34:15	U	SMPL	
3038-15 - 1	SB-09-COMP	14	0.0193	-7470A	PPB	16 Jul 2021 10:36:32	U	SMPL	
3038-16 - 1	SB-11-COMP	40	0.0261	-7470A	PPB	16 Jul 2021 10:38:48	U	SMPL	
3038-17 - 1	SB-10-COMP	35	0.0248	-7470A	PPB	16 Jul 2021 10:41:04	U	SMPL	
CV76 - 1	CCV76	18956	4.9884	-7470A	PPB	16 Jul 2021 10:43:19	U	SMPL	
CB76 - 1	CCB76	-82	-0.0059	-7470A	PPB	16 Jul 2021 10:45:34	U	SMPL	
3039-15 - 1	SB-12-COMP	19	0.0206	-7470A	PPB	16 Jul 2021 10:47:48	U	SMPL	
3039-16 - 1	SB-01-COMP	24	0.0219	-7470A	PPB	16 Jul 2021 10:50:04	U	SMPL	
3039-17 - 1	SB-02-COMP	40	0.0261	-7470A	PPB	16 Jul 2021 10:52:18	U	SMPL	
3039-18 - 1	SB-03-COMP	39	0.0259	-7470A	PPB	16 Jul 2021 10:54:34	U	SMPL	
3039-19 - 1	SB-04-COMP	45	0.0275	-7470A	PPB	16 Jul 2021 10:56:49	U	SMPL	
3039-20 - 1	SB-05-COMP	33	0.0243	-7470A	PPB	16 Jul 2021 10:59:05	U	SMPL	
3039-21 - 1	SB-08-COMP	42	0.0267	-7470A	PPB	16 Jul 2021 11:01:21	U	SMPL	
3039-21DUP - 1	SB-08-COMPDUP	31	0.0238	-7470A	PPB	16 Jul 2021 11:03:38	U	SMPL	
3039-21MS - 1	SB-08-COMPMS	14434	3.8021	-7470A	PPB	16 Jul 2021 11:05:55	U	SMPL	
3039-21MSD - 1	SB-08-COMPMSD	14615	3.8496	-7470A	PPB	16 Jul 2021 11:08:10	U	SMPL	
CV77 - 1	CCV77	19113	5.0296	-7470A	PPB	16 Jul 2021 11:10:25	U	SMPL	
CB77 - 1	CCB77	-93	-0.0087	-7470A	PPB	16 Jul 2021 11:12:41	U	SMPL	
B137717TB - 1	PB137701TB	-1	0.0154	-7470A	PPB	16 Jul 2021 11:14:56	U	SMPL	
3039-21LX5 - 1		25	0.0222	-7470A	PPB	16 Jul 2021 11:17:11	U	SMPL	
3039-21A - 1		14920	3.9296	-7470A	PPB	16 Jul 2021 11:19:26	U	SMPL	
CV78 - 1	CCV78	18530	4.8766	-7470A	PPB	16 Jul 2021 11:21:44	U	SMPL	
CB78 - 1	CCB78	-77	-0.0045	-7470A	PPB	16 Jul 2021 11:23:59	U	SMPL	