



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB115382

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/16/2021	11:45
2	CAL2	1	Water	NA	NA	21.2	7.00	07/16/2021	11:46
3	CAL3	1	Water	NA	NA	21.2	10.03	07/16/2021	11:47
4	ICV	1	Water	NA	NA	21.3	7.00	07/16/2021	11:49
5	CCV1	1	Water	NA	NA	20.2	2.01	07/16/2021	11:50
6	M3086-01	1	Water	NA	NA	20.1	4.66	07/16/2021	12:00
7	M3086-01DUP	1	Water	NA	NA	20.2	4.67	07/16/2021	12:01
8	M3086-02	1	Water	NA	NA	20.2	4.73	07/16/2021	12:03
9	M3086-03	1	Water	NA	NA	20.3	6.12	07/16/2021	12:05
10	CCV2	1	Water	NA	NA	21.3	12.02	07/16/2021	12:06

WORKLIST(Hardcopy Internal Chain)

LB 115382

WorkList Name : PH M3086 WorkList ID : 150632 Department : Wet-Chemistry Date : 07-16-2021 11:40:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/06/2021	Water	M3086-01	pH	Cool 4 deg C	STER09	L11	CELL-A-POOL-WATER(D	07/06/2021	9040C
07/06/2021	Water	M3086-02	pH	Cool 4 deg C	STER09	L11	CELL-B-C-POOL-WATER	07/06/2021	9040C
07/06/2021	Water	M3086-03	pH	Cool 4 deg C	STER09	L11	CELL-D-POOL-WATER(D	07/06/2021	9040C

Date/Time 07/16/21 11:42
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JPR (SM)

Date/Time 07/16/21 12:30
 Raw Sample Received by: JPR (SM)
 Raw Sample Relinquished by: JP (WC)