


LB115394

SFAM01.1

INSTRUMENT ID: CY1

 μ Abs.:

43915



Linear

A= 0.0000e+000

B= 2.2896e-004 *slope*C= -8.4105e-002 *y-intercept*

Rho= 0.9999868

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%D</i>
0.00	0.000	0.000	0.000	369	0.000	369					-6
0.05	0.050	0.047	-0.003	573	0.0 %	573					-1
0.20	0.200	0.198	-0.002	1234	0.0 %	1234					-1
2.50	2.500	2.483	-0.017	11213	0.0 %	11213					0
5.00	5.000	5.017	0.017	22281	0.0 %	22281					0
7.50	7.500	7.533	0.033	33268	0.0 %	33268					0
10.0	10.000	9.971	-0.029	43915	0.0 %	43915					0

LB115394
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	369	-		0 SFAM01.1	PPB	7/16/2021 14:51	Std
	0.05 S01	573	-	0.05	SFAM01.1	PPB	7/16/2021 14:54	Std
	0.2 S02	1234	-	0.2	SFAM01.1	PPB	7/16/2021 14:56	Std
	2.5 S03	11213	-	2.5	SFAM01.1	PPB	7/16/2021 14:58	Std
	5 S04	22281	-	5	SFAM01.1	PPB	7/16/2021 15:00	Std
	7.5 S05	33268	-	7.5	SFAM01.1	PPB	7/16/2021 15:03	Std
	10 S06	43915	-	10	SFAM01.1	PPB	7/16/2021 15:05	Std
ICV050	ICV050	17480	3.9181 -		SFAM01.1	PPB	7/16/2021 15:08	SMPL
ICB050	ICB050	323	-0.0102 -		SFAM01.1	PPB	7/16/2021 15:13	SMPL
CCV082	CCV082	22110	4.9782 -		SFAM01.1	PPB	7/16/2021 15:15	SMPL
CCB082	CCB082	305	-0.0143 -		SFAM01.1	PPB	7/16/2021 15:17	SMPL
PB137765BL	PBW765	337	-0.0069 -		SFAM01.1	PPB	7/16/2021 15:26	SMPL
M2976-01	EW125	536	0.0386 -		SFAM01.1	PPB	7/16/2021 15:28	SMPL
M2976-02	EW126	293	-0.017 -		SFAM01.1	PPB	7/16/2021 15:30	SMPL
M2976-03	EW126D	345	-0.0051 -		SFAM01.1	PPB	7/16/2021 15:32	SMPL
M2976-04	EW126S	4876	1.0323 -		SFAM01.1	PPB	7/16/2021 15:35	SMPL
CCV083	CCV083	21644	4.8715 -		SFAM01.1	PPB	7/16/2021 15:37	SMPL
CCB083	CCB083	288	-0.0182 -		SFAM01.1	PPB	7/16/2021 15:42	SMPL