



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB115400

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	07/16/2021	15:15
2	CAL2	1	Water	NA	NA	21.4	7.00	07/16/2021	15:16
3	CAL3	1	Water	NA	NA	21.4	10.03	07/16/2021	15:17
4	ICV	1	Water	NA	NA	21.4	7.00	07/16/2021	15:19
5	CCV1	1	Water	NA	NA	21.5	2.01	07/16/2021	15:20
6	M3061-01	1	Solid	20.02	20	22.6	11.08	07/16/2021	15:22
7	M3061-01DUP	1	Solid	20.03	20	22.7	11.09	07/16/2021	15:23
8	CCV2	1	Water	NA	NA	21.3	12.03	07/16/2021	15:25

2B 115400

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	CORROSIVITY M3061	WorkList ID :	150652	Department :	Wet-Chemistry	Date :	07-16-2021 15:05:52		
Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/14/2021	Solid	M3061-01	Corrosivity	Cool 4 deg C	ARCA05	L11	PB-2021-WW-000-08	07/14/2021	9045D

Date/Time 07-16-21 15:16
Raw Sample Received by: JA (WC)
Raw Sample Relinquished by: JA (WC)

Date/Time 07-16-21 16:10
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)