



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 7/20/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 07/19/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:19
Out Date: 07/20/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB115416

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M3070-01	01-A-01-B-01-C	1	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3070-02	02-A-02-B-02-C	2	1.00	2.00	2.00	100.0	CAULKING SAMPLE
M3088-01	B1- (A-D) - (0-4-2)	3	1.12	9.87	8.85	88.3	
M3088-03	B1-D-BOTTOM- (4-2)	4	1.12	9.9	8.84	87.9	
M3089-01	PAR-N2- (A-D) - (0-12)	5	1.16	9.86	8.86	88.5	
M3089-03	PAR-N2-D-BOTTOM- (4-11)	6	1.13	9.84	8.85	88.6	
M3089-12	R99997- (A-D) - (0-3)	7	1.15	9.95	8.43	82.7	
M3089-14	R99997-D-BOTTOM- (3)	8	1.19	9.98	8.2	79.7	
M3093-01	20210528-COM-1	9	1.12	9.68	7.05	69.3	sediment sample
M3093-02	20210529-COM-2	10	1.16	9.54	4.49	39.7	sediment sample
M3093-03	20210531-COM-4	11	1.15	9.91	5.85	53.7	sediment sample
M3093-04	20210537-COM-10	12	1.15	9.96	7.22	68.9	sediment sample
M3093-05	20210543-COM-16	13	1.16	9.96	5.63	50.8	sediment sample
M3093-06	20210557-COM-30	14	1.11	9.88	5.4	48.9	sediment sample
M3094-01	SP-1	15	1.12	9.62	8.61	88.1	
M3094-01D	SP-1DUP	16	1.19	9.64	8.63	88.0	
M3094-02	SP-1A	17	1.14	9.97	8.75	86.2	
M3095-01	H151-I1-001	22	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3095-02	H151-I2-001	23	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3095-03	H151-O1-001	24	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3095-04	H151-O2-001	25	1.00	2.00	2.00	100.0	WIPE SAMPLE
M3096-01	SU-04-071921	18	1.14	9.83	8.88	89.1	
M3096-02	SU-04-071921-EPH-2	19	1.12	9.59	8.72	89.7	
M3097-01	HD-01-071921	20	1.15	9.97	9.68	96.7	
M3097-02	HD-01-071921-EPH-2	21	1.11	9.59	9.44	98.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$