

7471B


LB115419

INSTRUMENT ID: CV2

Linear

 μ Abs.:

48704



A= 0.0000e+000

B= 2.0473e-004

C= -4.8140e-002

Rho= 0.9997414

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.042	-0.042	32	0.000	32				5
0.2	0.200	0.209	0.009	1258	0.0 %	1258				4
2.5	2.500	2.432	-0.068	12115	0.0 %	12115				-3
5.0	5.000	5.173	0.173	25502	0.0 %	25502				3
7.5	7.500	7.504	0.004	36888	0.0 %	36888				0
10.0	10.000	9.923	-0.077	48704	0.0 %	48704				-1

LB115419 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 19 Jul 2021 14:57:20				
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type
0.0 - 1	50	32	-	0.0000	7471B	PPB	19 Jul 2021 15:01:54	S Std
0.2 - 1	50-2	1258	-	0.2000	7471B	PPB	19 Jul 2021 15:04:10	S Std
2.5 - 1	52-5	12115	-	2.5000	7471B	PPB	19 Jul 2021 15:06:28	S Std
5.0 - 1	53	25502	-	5.0000	7471B	PPB	19 Jul 2021 15:11:23	S Std
7.5 - 1	53-5	36888	-	7.5000	7471B	PPB	19 Jul 2021 15:16:09	S Std
10.0 - 1	54	48704	-	10.0000	7471B	PPB	19 Jul 2021 15:18:25	S Std
ICV11 - 1	ICV11	19439	3.9316	-	7471B	PPB	19 Jul 2021 15:23:40	U SMPL
ICB11 - 1	ICB11	125	-0.0225	-	7471B	PPB	19 Jul 2021 15:25:54	U SMPL
CCV87 - 1	CCV87	26166	5.3089	-	7471B	PPB	19 Jul 2021 15:28:09	U SMPL
CCB87 - 1	CCB87	117	-0.0242	-	7471B	PPB	19 Jul 2021 15:30:24	U SMPL
CRA - 1	CRA	1247	0.2072	-	7471B	PPB	19 Jul 2021 15:32:40	U SMPL
HighStd - 1	HighStd	49036	9.9911	-	7471B	PPB	19 Jul 2021 15:34:55	U SMPL
ChkStd - 1	ChkStd	35037	7.1250	-	7471B	PPB	19 Jul 2021 15:37:09	U SMPL
PB137794BL - 1	PBS	193	-0.0086	-	7471B	PPB	19 Jul 2021 15:44:27	U SMPL
PB137794BS - 1	LCSS	20777	4.2056	-	7471B	PPB	19 Jul 2021 15:46:42	U SMPL
M3069-01 - 1	CL-01-071621	3551	0.6789	-	7471B	PPB	19 Jul 2021 15:48:58	U SMPL
M3072-01 - 1	OR-02-071621	1825	0.3255	-	7471B	PPB	19 Jul 2021 15:51:15	U SMPL
M3074-01 - 1	OK-02-071621	1363	0.2309	-	7471B	PPB	19 Jul 2021 15:53:32	U SMPL
M3076-01 - 1	20210578-COM-1	27334	5.5480	-	7471B	PPB	19 Jul 2021 15:55:49	U SMPL
M3076-02 - 1	20210587-AMENDED-COM-1	25375	5.1469	-	7471B	PPB	19 Jul 2021 15:58:04	U SMPL
CCV88 - 1	CCV88	25396	5.1512	-	7471B	PPB	19 Jul 2021 16:00:20	U SMPL
CCB88 - 1	CCB88	194	-0.0084	-	7471B	PPB	19 Jul 2021 16:05:06	U SMPL
M3076-03 - 1	20210534-COM-7	42300	8.6120	-	7471B	PPB	19 Jul 2021 16:07:20	U SMPL
M3083-01 - 1	TT-087-RW7-IDWSO-20210715	763	0.1081	-	7471B	PPB	19 Jul 2021 16:09:35	U SMPL
M3083-01DUP - 1	TT-087-RW7-IDWSO-20210715DUP	756	0.1066	-	7471B	PPB	19 Jul 2021 16:11:53	U SMPL
M3083-01MS - 1	TT-087-RW7-IDWSO-20210715MS	20650	4.1796	-	7471B	PPB	19 Jul 2021 16:14:08	U SMPL
M3083-01MSD - 1	TT-087-RW7-IDWSO-20210715MSD	20866	4.2238	-	7471B	PPB	19 Jul 2021 16:16:24	U SMPL
M3088-01 - 1	B1-(A-D)-(0-4-2)	2726	0.5100	-	7471B	PPB	19 Jul 2021 16:18:40	U SMPL
M3089-01 - 1	PAR-N2-(A-D)-(0-12)	4259	0.8238	-	7471B	PPB	19 Jul 2021 16:20:57	U SMPL
M3089-12 - 1	R99997-(A-D)-(0-3)	10279	2.0563	-	7471B	PPB	19 Jul 2021 16:23:13	U SMPL
M3083-01LX5 - 1		203	-0.0066	-	7471B	PPB	19 Jul 2021 16:25:31	U SMPL
M3083-01A - 1		20971	4.2453	-	7471B	PPB	19 Jul 2021 16:27:46	U SMPL
CCV89 - 1	CCV89	25481	5.1686	-	7471B	PPB	19 Jul 2021 16:30:01	U SMPL
CCB89 - 1	CCB89	98	-0.0281	-	7471B	PPB	19 Jul 2021 16:32:17	U SMPL