



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115421

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	07/19/2021	14:05
2	M3039-22		1	9	100.00	>212.0	>212.0	07/19/2021	14:35
3	M3039-22DUP		1	9	100.00	>212.0	>212.0	07/19/2021	14:58
4	M3051-09		1	8	100.00	>212.0	>212.0	07/19/2021	15:20
5	M3051-10		1	8	100.00	>212.0	>212.0	07/19/2021	15:45
6	M3051-11		1	9	100.00	>212.0	>212.0	07/19/2021	16:10
7	M3061-02		1	9	100.00	>212.0	>212.0	07/19/2021	16:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP71921-1 WorkList ID : 150693 Department : Wet-Chemistry Date : 07-19-2021 13:32:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/21/2021	Water	M3039-22	Flash Point	Cool 4 deg C	L1RO01	L31	EX-05	07/14/2021	1010B
07/22/2021	Water	M3051-09	Flash Point	Cool 4 deg C	PSEG03	L33	41213	07/15/2021	1010B
07/22/2021	Water	M3051-10	Flash Point	Cool 4 deg C	PSEG03	L33	COMP-41257-59	07/15/2021	1010B
07/22/2021	Water	M3051-11	Flash Point	Cool 4 deg C	PSEG03	L33	30817	07/15/2021	1010B
07/22/2021	Water	M3061-02	Flash Point	Cool 4 deg C	ARCA05	L11	PB-2021-WW-000-08	07/14/2021	1010B

Date/Time 07/19/21 13:40
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: SP (WC)

Date/Time 07/19/21 16:45
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: KS (WC)