



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB115424

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP92115
Turbidity Calibration std, 0NTU	WP92109
Turbidity Calibration std, 20NTU	WP92111
Turbidity Calibration std, 40NTU	WP92110
Turbidity Calibration std, 1NTU	WP92114
Turbidity Calibration std, 5NTU	WP92113
Turbidity Calibration - CCV std, 10 NTU	WP92112
10 NTU Standard 500 ml	W2752

Intercept: 0.0933

Slope: 1.0102

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.185	0.09		07/19/2021	16:06
2	CAL2	1	1	1.180	1.08	-7.6	07/19/2021	16:08
3	CAL3	5	1	5.220	5.08	-1.5	07/19/2021	16:11
4	CAL4	10	1	9.950	9.76	2.4	07/19/2021	16:13
5	CAL5	20	1	20.200	19.90	0.5	07/19/2021	16:16
6	CAL6	40	1	40.600	40.10	-0.2	07/19/2021	16:19



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: apatel

Parameter: Turbidity

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB115424

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.400	10.203	07/19/2021	16:24
2	ICB		1	0.189	0.095	07/19/2021	16:26
3	CCV1	10	1	10.500	10.302	07/19/2021	16:28
4	CCB1		1	0.191	0.097	07/19/2021	16:30
5	LB115424BL		1	0.191	0.097	07/19/2021	16:33
6	M2940-12		1	0.552	0.454	07/19/2021	16:35
7	M3035-02		1	1.360	1.254	07/19/2021	16:38
8	M3035-02DUP		1	1.370	1.264	07/19/2021	16:41
9	CCV2	10	1	10.400	10.203	07/19/2021	16:44
10	CCB2		1	0.192	0.098	07/19/2021	16:49