



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB115438

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/20/2021	13:45
2	M3051-02		1	4	20.00	68.0	68.0	07/20/2021	14:10
3	M3051-02DUP		1	3	19.00	66.2	66.2	07/20/2021	14:32
4	M3051-04		1	5	60.00	140.0	140.0	07/20/2021	14:58
5	M3051-06		1	6	55.00	131.0	131.0	07/20/2021	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB115438

WorkList Name : FP71921 WorkList ID : 150692 Department : Wet-Chemistry Date : 07-19-2021 13:31:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/22/2021	Solid	M3051-02	Flash Point	Cool 4 deg C	PSEG03	L33	30778	07/15/2021	1010B
07/22/2021	Solid	M3051-04	Flash Point	Cool 4 deg C	PSEG03	L33	41904	07/15/2021	1010B
07/22/2021	Solid	M3051-06	Flash Point	Cool 4 deg C	PSEG03	L33	41272	07/15/2021	1010B

Date/Time 07/20/21 13:25
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CFA-1066

Date/Time 07/20/21 15:25
 Raw Sample Received by: RJ CFA-1066
 Raw Sample Relinquished by: KS (WC)