


LB115440

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9184e-004

C= -1.1283e-002

Rho= 0.9994606

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.074	0.074	442	0.000	442					6
0.2	0.200	0.213	0.013	1167	0.0 %	1167					-8
2.5	2.500	2.309	-0.191	12094	0.0 %	12094					4
5.0	5.000	5.185	0.185	27087	0.0 %	27087					-6
7.5	7.500	7.395	-0.105	38604	0.0 %	38604					0
10.0	10.000	10.025	0.025	52315	0.0 %	52315					

LB115440 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Jul 2021 15:35:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	442	-	0.0000	7471B	PPB	20 Jul 2021 15:38:17	S	Std
0.2 - 1	50.2	1167	-	0.2000	7471B	PPB	20 Jul 2021 15:40:34	S	Std
2.5 - 1	52.5	12094	-	2.5000	7471B	PPB	20 Jul 2021 15:42:52	S	Std
5.0 - 1	55	27087	-	5.0000	7471B	PPB	20 Jul 2021 15:45:08	S	Std
7.5 - 1	57.5	38604	-	7.5000	7471B	PPB	20 Jul 2021 15:50:00	S	Std
10.0 - 1	510	52315	-	10.0000	7471B	PPB	20 Jul 2021 15:52:15	S	Std
ICV13 - 1	ICV13	21146	4.0455	-	7471B	PPB	20 Jul 2021 15:57:30	U	SMPL
ICB13 - 1	ICB13	68	0.0018	-	7471B	PPB	20 Jul 2021 16:01:41	U	SMPL
CCV95 - 1	CCV95	27416	5.2483	-	7471B	PPB	20 Jul 2021 16:03:56	U	SMPL
CCB95 - 1	CCB95	-35	-0.0180	-	7471B	PPB	20 Jul 2021 16:06:11	U	SMPL
CRA - 1	CRA	1158	0.2109	-	7471B	PPB	20 Jul 2021 16:08:25	U	SMPL
HighStd - 1	HighStd	52957	10.1482	-	7471B	PPB	20 Jul 2021 16:10:40	U	SMPL
ChkStd - 1	ChkStd	37060	7.0985	-	7471B	PPB	20 Jul 2021 16:12:55	U	SMPL
PB137833BL - 1	LBS	-169	-0.0437	-	7471B	PPB	20 Jul 2021 16:15:13	U	SMPL
PB137833BS - 1	LCSS	19258	3.6833	-	7471B	PPB	20 Jul 2021 16:20:01	U	SMPL
M3093-01 - 1	20210528-COM-1	28990	5.5503	-	7471B	PPB	20 Jul 2021 16:22:17	U	SMPL
M3093-02 - 1	20210529-COM-2	24470	4.6832	-	7471B	PPB	20 Jul 2021 16:24:33	U	SMPL
M3093-03 - 1	20210531-COM-4	28975	5.5474	-	7471B	PPB	20 Jul 2021 16:26:51	U	SMPL
M3093-04 - 1	20210537-COM-10	47157	9.0355	-	7471B	PPB	20 Jul 2021 16:29:09	U	SMPL
M3093-05 - 1	20210543-COM-16	16390	3.1331	-	7471B	PPB	20 Jul 2021 16:31:24	U	SMPL
CCV96 - 1	CCV96	27453	5.2554	-	7471B	PPB	20 Jul 2021 16:33:40	U	SMPL
CCB96 - 1	CCB96	-41	-0.0191	-	7471B	PPB	20 Jul 2021 16:35:55	U	SMPL
M3093-06 - 1	20210557-COM-30	11600	2.2141	-	7471B	PPB	20 Jul 2021 16:38:11	U	SMPL
M3094-01 - 1	SP-1	1682	0.3114	-	7471B	PPB	20 Jul 2021 16:40:26	U	SMPL
M3096-01 - 1	SU-04-071921	3184	0.5996	-	7471B	PPB	20 Jul 2021 16:42:41	U	SMPL
M3096-01DUP - 1	SU-04-071921DUP	3396	0.6402	-	7471B	PPB	20 Jul 2021 16:47:38	U	SMPL
M3096-01MS - 1	SU-04-071921MS	21336	4.0819	-	7471B	PPB	20 Jul 2021 16:49:53	U	SMPL
M3096-01MSD - 1	SU-04-071921MSD	22633	4.3307	-	7471B	PPB	20 Jul 2021 16:52:10	U	SMPL
M3097-01 - 1	HD-01-071921	1809	0.3358	-	7471B	PPB	20 Jul 2021 16:54:26	U	SMPL
M3096-01LX5 - 1		770	0.1364	-	7471B	PPB	20 Jul 2021 16:56:43	U	SMPL
M3096-01A - 1		23889	4.5717	-	7471B	PPB	20 Jul 2021 16:58:59	U	SMPL
CCV97 - 1	CCV97	27621	5.2877	-	7471B	PPB	20 Jul 2021 17:01:15	U	SMPL
CCB97 - 1	CCB97	-25	-0.0161	-	7471B	PPB	20 Jul 2021 17:03:30	U	SMPL